

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal|sup|TS|sup0 Standard, 1000 NTU**COMMODITY NUMBER: **2660642-BR** MANUFACTURE DATE:LOT NUMBER: **A2034****4/3/2022**

DATE OF ANALYSIS:

4/3/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	950 to 1050 NTU	995.3 NTU

The expiration date is Jun 2024

Certified by

Scott Als
Analytical Services Chemist