

HACH COMPANY P.O.Box 389



Loveland, CO 80539
(970) 669-3050

An ISO 9001 Certified Company

Certificate of Analysis

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD** 20 NTU

COMMODITY NUMBER: **2684801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A9188**

7/4/2020

7/11/2020

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	19 to 21 NTU	20.3 NTU

The expiration date is Jul 2021

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26848.01</u>
Data: <u>17/08/2020</u>
Validade: <u>01/07/2021</u>

Certified by _____

A handwritten signature in cursive script that reads "Scott Als".

Scott Als
Analytical Services Chemist