

Certificate of Analysis List

For request number 1049077

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
29153 as F	6236	N/A	N/A	Fluoride Standard Solution 1.00 mg/l

Total Enclosures: 1

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **Fluoride Standard Solution 1.00 mg/l as F**

COMMODITY NUMBER: **29153**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A6236**

8/22/2016

8/22/2016

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Fluoride concentration by SPADNS method.	0.98 to 1.02 ppm	0.980 ppm

The expiration date is Aug 2021

The item 29153 is traceable to NIST standards SRM 3183 Fluoride 1 mg/mL LOT 050721.

Certified by _____

Scott Als
Analytical Services Chemist