



An ISO 9001 Certified Company

**Certificate of Analysis****This is a Component of 2971205 lot A10032**

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 800 NTU**COMMODITY NUMBER: **2660501**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: A10015

**4/23/2022****4/31/2022**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	760 to 840 NTU	812.6 NTU

The expiration date is Apr 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

<b>APROVADO</b>
Responsável: <u>RENATO</u>
Padrão: <u>26605.01</u>
Data: <u>02/06/2022</u>
Validade: <u>30/04/2023</u>

Certified by

Scott Als  
Analytical Services Chemist