

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389  
Loveland, CO 80539  
(970) 669-3050

*Certificate of Analysis*

Page 1

COMMODITY: **10 NTU Verification Standard**

COMMODITY NUMBER: **2961801**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A7345A**

**12/12/2017**

**12/19/2017**

---

<i><b>TEST</b></i>	<i><b>SPECIFICATIONS</b></i>	<i><b>RESULTS</b></i>
Turbidity	9.5 to 10.5 NTU	10.22 NTU

The expiration date is Mar 2019

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

A handwritten signature in cursive script that reads "Scott Als".

Certified by \_\_\_\_\_

Scott Als  
Analytical Services Chemist