

***Certificate of Analysis******This is a Component of 2659505-BR lot A1355***

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD 20 NTU**COMMODITY NUMBER: **2660101**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A1320****12/1/2021****12/7/2021**

---

<b>TEST</b>	<b>SPECIFICATIONS</b>	<b>RESULTS</b>
Turbidity	19 to 21 NTU	20.5 NTU

The expiration date is Mar 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIS does not carry turbidity standards. However, the use of Formazin and StablCal® as use in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

<b>APROVADO</b>
Responsável: <u>RENATO</u>
Padrão: <u>26601.01</u>
Data: <u>02/06/2022</u>
Validade: <u>30/03/2023</u>

Certified by

Scott Als  
Analytical Services Chemist