

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2971205-BR lot A1027

Page 1

COMMODITY: **STABLCAL® FORMAZIN STANDARD 100 NTU**
COMMODITY NUMBER: **2684901** MANUFACTURE DATE:
LOT NUMBER: **A1032** **2/9/2021**

DATE OF ANALYSIS:
2/12/2021

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	95 to 105 NTU	103.0 NTU

The expiration date is May 2022

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO	
Responsável:	<u>Wellington</u>
Ideração:	<u>2684901-CJS</u>
Data:	<u>06/07/2021</u>
Validade:	<u>30/05/2022</u>

Certified by

Scott Als

Scott Als
Analytical Services Chemist