

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

This is a Component of 2971205-BR lot A2180

Page 1

COMMODITY: **STABLCAL|sup|TS|sup0 FORMAZIN STANDARD** 20 NTU
COMMODITY NUMBER: **2684801** MANUFACTURE DATE: DATE OF ANALYSIS:
LOT NUMBER: **A2112** 5/3/2022 5/6/2022

TEST	SPECIFICATIONS	RESULTS
Turbidity	19 to 21 NTU	20.1 NTU

The expiration date is Jul 2023

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

APROVADO
Responsável: <u>RENATO</u>
Padrão: <u>26848.01</u>
Data: <u>13/10/2022</u>
Validade: <u>30/07/2023</u>

Certified by _____

Scott Als
Analytical Services Chemist