

Certificate of Analysis List

For request number 828007

Catalog Number Entered	Lot Number Entered	Related Catalog Number	Related Lot Code	Description
2659942	6207	N/A	N/A	StablCal [®] Standard, 10 NTU

Total Enclosures: 1

*Certificate of Analysis*

Page 1

COMMODITY: **StablCal® Standard, 10 NTU**COMMODITY NUMBER: **26599-42**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A6207****7/24/2016****7/27/2016**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	10.20 NTU

The expiration date is Jul 2018

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist