

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **StablCal® Standard, 10 NTU**

COMMODITY NUMBER: **2659942**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A7209**

8/4/2017

8/7/2017

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Turbidity	9.5 to 10.5 NTU	10.10 NTU

The expiration date is Jul 2019

Formazin and StablCal® solutions provided by Hach are not NIST traceable because the NIST does not carry turbidity standards. However, the use of Formazin and StablCal® as used in Hach method 8195 are accepted by the EPA as a primary standard to be used in the calibration of turbidity instruments.

Certified by _____

Scott Als
Analytical Services Chemist