

Problem scale: original 2D GCNM vs subject-6 PVI

Quantity	Original (Injun)	Subject-6 PVI
Physical electrodes	32	8
Measurements per frame	928	32
Inverse elements, K	—	1,330
Forward elements	—	5,320
Measurement-to-unknown ratio	high	$32/1330 \approx 0.024$

29× fewer boundary measurements for a larger unknown field.

Takeaway. The PVI ring supplies 29× fewer measurements for a larger unknown field. This information deficit is fundamental—no amount of network capacity removes it.

Source: technical report Tables 1-2 and Figure 1.