

CBCT Modes TrueBeam 4.1

Date printed: 30/07/2025

Mode Name	Dose (mSv)	Base mode	Trajectory	kV	mAs Per frame	frame/s	deg/s	Fan	# Proj	Acq time	Exposure	CTDI
4D Spotlight	1.6	Standard	Half	125	0.8	7	3	Full	467	67	373.6	6.1
4D Thorax	3.3	Standard	Full	125	0.8	7	3	Half (LFOV)	840	120	671.2	9.9
Abdo Spotlight	1.2	Spotlight	Half	125	0.8	15	6	Full (SFOV)	500	33	400.8	6.6
Abdomen	2.8	Thorax	Full	125	0.8	15	6	Half (LFOV)	900	60	716	10.6
Breast 360	0.2	Thorax	Full	125	0.1	15	6	Half (LFOV)	900	60	89.5	1.3
Extremity Spotlight	0.5	Head	Half	100	0.3	15	6	Full (SFOV)	500	33	150.3	3.2
Head	0.5	Standard	Half	100	0.3	15	6	Full (SFOV)	500	33	150.3	3.2
Head and Shoulders	0.3	Thorax	Full	125	0.3	15	6	Half (LFOV)	900	60	268.5	4.0
Head SRS	1.8	Head	Full	100	0.6	15	6	Full (SFOV)	900	60	537	11.3
Image Gently		Standard	Half	80	0.2	15	6	Full (SFOV)	500	33	100.2	0.9
Paediatric Body		Image Gently	Half	80	0.2	15	6	Full (SFOV)	500	33	100.2	0.9
Pediatric Head		Image Gently	Half	80	0.1	15	6	Full (SFOV)	500	33	50.1	0.5
Pelvis	4.2	Standard	Full	125	1.2	15	6	Half (LFOV)	900	60	1074	15.9
Pelvis Large		Pelvis Obese	Full	140	1.9	15	6	Half (LFOV)	900	60	1700.5	37.1
Pelvis Spotlight	2.2	Spotlight	Half	125	1.5	15	6	Full (SFOV)	500	33	751.5	12.3
SBRT Spine	1.7	Thorax	Full	125	0.4	15	6	Half (LFOV)	900	60	358	5.3
Thorax	1.3	Standard	Full	125	0.3	15	6	Half (LFOV)	900	60	268.5	4.0
Thorax Spotlight	0.7	Spotlight	Half	125	0.3	15	6	Full (SFOV)	500	33	150.3	2.5

Bold indicates change from Varian standard settings

Effective Dose mSv estimates from PCXMC calculations for generic 80kg/180cm person, scales with body habitus

TB 4.1 upgrade combines mA & S single value, CTDI now in mGy

Documents - ROMP-WSLHD\03 Projects\2023 - CBCT Mode Standardisation