

CBCT Modes

Current

As of: 24/04/2024

		Acquisition											
		/mSv											
Mode Name	Dose	Base mode	Trajectory	kV	mA/f	ms/f	f/s	deg/s	Fan	No Proj	Acq time	Exposure	CTDI
4D Spotlight	1.6	Standard	Half	125	40	20	7	3	Full	467	67	374	0.61
4D Thorax	3.3	Standard	Full	125	40	20	7	3	Half (LFOV)	840	120	672	1.00
Abdomen	2.8	Thorax	Full	125	40	20	15	6	Half (LFOV)	900	60	720	1.07
Abdo Spotlight	1.2	Spotlight	Half	125	40	20	15	6	Full (SFOV)	500	33	400	0.66
Breast 360	0.2	Thorax	Full	125	10	10	15	6	Half (LFOV)	900	60	90	0.13
Head	0.5	Standard	Half	100	15	20	15	6	Full (SFOV)	500	33	150	0.32
Head SRS	1.8	Head	Full	100	30	20	15	6	Full (SFOV)	900	60	540	1.14
Head and Shoulders	0.3	Thorax	Full	125	15	20	15	6	Half (LFOV)	900	60	270	0.40
Image Gently		Standard	Half	80	20	10	15	6	Full (SFOV)	500	33	100	0.09
Pelvis	4.2	Standard	Full	125	60	20	15	6	Half (LFOV)	900	60	1080	1.60
Pelvis Large		Pelvis Obese	Full	140	75	25	15	6	Half (LFOV)	900	60	1688	3.68
Pelvis Spotlight	2.2	Spotlight	Half	125	60	25	15	6	Full (SFOV)	500	33	750	1.23
Thorax	1.3	Standard	Full	125	15	20	15	6	Half (LFOV)	900	60	270	0.40
Thorax Spotlight	0.7	Spotlight	Half	125	15	20	15	6	Full (SFOV)	500	33	150	0.20
Extremity Spotlight	0.5	Head	Half	100	15	20	15	6	Full (SFOV)	500	33	150	0.32
SBRT Spine	1.7	Thorax	Full	125	40	10	15	6	Half (LFOV)	900	60	360	0.53
Pediatric Head		Image Gently	Half	80	10	10	15	6	Full (SFOV)	500	33	50	0.05
Paediatric Body		Image Gently	Half	80	20	10	15	6	Full (SFOV)	500	33	100	0.09

Nonclinical CBCT modes are not on this list (e.g., factory defaults).

Bold indicates change from Varian standard settings

Effective Dose mSv estimates from PCXMC calculations for generic 80kg/180cm person, scales with body habitus