

CAUTION: TO PREVENT INJURY AND DAMAGE TO EQUIPMENT, THE OPERATION AND/OR MAINTENANCE SHOULD BE CARRIED OUT BY A QUALIFIED TECHNICIAN.

Item	Before Each Test	✓
1	Inspect the temperature measurement system.	
2	Perform a self-calibration of the temperature scanners.	

Item	1 Month	✓
1	Perform a self-calibration of the temperature scanners.	

Item	6 Months	✓
1	Inspect the temperature measurement system.	

Item	Annual	✓
1	Audit the analogue temperature channels.	
2	Inspect the calibration of a temperature scanner.	

Item	As Required	✓
1	Calibrate the analogue temperature channels.	
2	Troubleshooting trigger problems with temperature scanners.	

Comments:

Checklist Completed By:	Date