Physics-Calibrated Simulation of Deformable Tissue for Ultrasound-Guided Robotic Femoral Access

Amit Manicka[1], Laurent Coulon[2], Darwin Ang[2], Michael Loushin[3] and Timothy M. Kowalewski[1]

***Abstract*— In ultrasound-guided femoral access, probe pressure deforms the tissue and changes the vascular anatomy the robot observes. A simulator must reproduce these mechanics fast enough for robotic interaction, yet the computational effort needed for a clinically meaningful signal is poorly characterized. We test the feasibility of a physics-based simulation of ultrasound-guided femoral access in Isaac Sim. The tissue is modeled with a Stable Neo-Hookean constitutive model and solved with Vertex Block Descent on one GPU. A simulation of a low-fidelity benchtop phantom pressed by a robot predicted its artery within 5 percentage points at every depth. It matched the vein within 3 points through 11 mm of a 25.4 mm press, then collapsed it more than the benchtop phantom did. We then map how element size, time step and solver iterations trade deformation accuracy against speed, with three to five runs per setting. Wall clock per frame scaled linearly with solver work, but accuracy did not improve with more work. The chosen setting cut the solver's time per frame from 29,700 ms to 30.4 ms, while vein and artery lumen areas stayed within 1.03 percentage points of the finest reference setting. With ultrasound and a rendered scene, the full loop took 56.4 ms per frame. Given literature-based tissue properties, the leg model compressed the vein far more than the artery at every solver setting. This reproduces the compressibility criterion clinicians use to distinguish them.**

I. Introduction

Simulation is now a standard environment for training and evaluating medical robots [1]–[3], but the mechanics of current robot-learning frameworks [4], [5] remain strongest for rigid bodies [6], [7], and probe-control policies for robotic ultrasound are typically trained on static image volumes or rigid scenes [1], [2], [8], [9]. Deformable solvers exist, but their calibration against real tissue is rarely reported [7], [10], [11], so a simulator can look anatomically correct while behaving mechanically inaccurately.

This matters because probe pressure, needle insertion and vessel compression all change the anatomy the robot observes [1], [12], and the ultrasound image measures the deformed anatomy rather than a fixed model [13], [14]. The chain runs from a robot action to a contact force, to the deformed tissue, to the image that drives the next action (Fig. S1). Sim-to-real transfer depends more on observations that stay consistent with probe motion than on single-frame realism [15].

We target fully automated access to both the common femoral vein and artery for emergency medicine and resuscitation. In these settings, aortic balloon occlusion needs arterial access [16], extracorporeal life support can need both vessels [17], and an absent pulse makes ultrasound replace palpation [18], [19]. Probe pressure collapses the thin-walled vein while the artery holds its lumen [20], and clinical guidance uses this compressibility to tell them apart [21], [22]. Although color Doppler can also separate them [18], we do not assume it is available in emergency use, and thus rely on the B-mode image. Most robotic ultrasound systems regulate contact force to suppress probe-induced deformation [13], [23]–[25]. Only one has used compression to classify the vessels [26]. Compression also has a cost: a compressed vein is a smaller target, so clinicians press lightly during puncture [27]. We envision a brief, deeper press that labels the vessels, then a lighter contact that keeps the vein open for the needle. However, both need a model that predicts vein and artery lumen area at each probe depth. In the Seldinger technique, the needle, dilator and catheter each compress the vein again, and every later step inherits the puncture site chosen under the probe [28]. This work therefore starts with probe contact.

A useful simulator must reproduce this behavior fast enough for a robot to interact with it, and for the thousands of repetitions that learning demands [2], [4]. We report cost as wall-clock milliseconds per frame, where a frame is 16.7 ms of simulated time, one image at the 60 Hz rate of an ultrasound display. Speed is not the only concern: a numerically accurate solution can still rest on material parameters that do not represent real tissue [6], [11]. *This work therefore asks how much computational effort is required to reproduce a clinical signal faithfully enough for a robot to interact with?* We test whether Isaac Sim is feasible for this task. We build an interactive Stable Neo-Hookean model [29] of the femoral region, solved on the GPU with Vertex Block Descent [30], whose deformation drives the rendered ultrasound image. We first compare a simulation of a simple benchtop phantom of known materials with robot presses on it. We then map how element size, time step and iterations trade accuracy against speed, first on a tissue slab, then on the leg. Finally, we give the leg model literature-based tissue properties and test whether it separates the femoral vein from the artery. Supplementary material (Sections S1–S3, Figs. S1–S7, Tables S1–S9, videos, data, scripts) is available upon request.

[1]Department of Mechanical Engineering, University of Minnesota, Minneapolis, MN, USA.
[2]Lightside Surgical, Seattle, WA, USA.
[3]Visible Heart Laboratory, Department of Anesthesiology, University of Minnesota Medical School, Minneapolis, MN, USA.

II. Methods

A. Units and Names

Simulated time passes inside the model while *wall-clock time* is what the computer takes. Frames advance in time steps Δt, often called substeps elsewhere. Lumen area is a percent opening of the same lumen at rest: differences in it are percentage points ("points"). A depth is *settled* when lumen area and displacement vary by under 2% over 30 frames (500 ms of simulated time). The *leg model* is the simulated anatomy. The *benchtop phantom* is the physical gel phantom, and the *simulated phantom* is its model. The *reference setting* is the finest leg setting completed, and the *chosen setting* is the fast setting selected for use.

B. Deformable Tissue Simulation

The tissue is a tetrahedral finite-element mesh with the Stable Neo-Hookean constitutive model [29]. It is solved with Vertex Block Descent (VBD) [30], a solver from the computer graphics literature whose per-vertex updates parallelize well on a GPU (Fig. S2). Contact enters as penalty energies in the same per-vertex minimization rather than as a separate constraint solve. Each tetrahedron has a deformation gradient $\mathbf{F} = \mathbf{D}_s\mathbf{D}_m^{-1}$, where $\mathbf{D}_s$ and $\mathbf{D}_m$ hold the current and rest edge vectors from vertex 0 [31]. Its strain energy density is

$$\Psi = \frac{\hat{\mu}}{2}(I_C - 3) + \frac{\hat{\lambda}}{2}(J - \alpha)^2, \tag{1}$$

with $I_C = \|\mathbf{F}\|_F^2$, $J = \det \mathbf{F}$ the local volume ratio, $\alpha = 1 + \hat{\mu}/\hat{\lambda}$, $\hat{\mu} = \mu$ and $\hat{\lambda} = \lambda + \mu$, where the Lamé parameters $\mu = E/[2(1+\nu)]$ and $\lambda = E\nu/[(1+\nu)(1-2\nu)]$ follow from Young's modulus E and Poisson ratio ν (Table II), so the energy matches linear elasticity at rest [29]. A tetrahedron straddling two materials takes the mean of its sub-samples, excluding bone. Inverted elements ($J < 0$) keep a defined energy [29], so the solver continues past them. At each time step VBD minimizes

$$\mathcal{E}(\mathbf{x}) = \sum_i \frac{m_i}{2\Delta t^2}\,\|\mathbf{x}_i - \tilde{\mathbf{x}}_i\|^2 + \sum_e V_e\,\Psi_e(\mathbf{x}) + \mathcal{E}_{\text{contact}}(\mathbf{x}), \tag{2}$$

where m_i is the mass of vertex i, $\tilde{\mathbf{x}}_i$ its inertial target, V_e and Ψ_e the rest volume and energy density of element e, and $\mathcal{E}_{\text{contact}}$ the probe and vessel-wall contact energies (Section II-C). VBD moves one vertex at a time with the others fixed, by a Newton step on the 3×3 block Hessian of the elements and contacts touching that vertex. Vertices sharing no element are independent. A graph coloring groups them into colors, each color is updated in parallel, and one iteration is one pass over all colors. Iterations per step are a fixed count, with no error tolerance and no subdivision of the time step. The solver (Newton 1.2.1) adds a rate-dependent damping stress with $\gamma = 1.0$ s in every element, which vanishes at rest (Section S1).

In each frame, the deformed tissue surfaces are passed to raysim, the ultrasound simulator of NVIDIA's Isaac for Healthcare [5], which ray traces a B-mode image from them. Isaac Sim renders the probe and the leg in the same loop. All runs used one workstation with an NVIDIA GeForce RTX 5090 GPU (32 GB), an Intel Core Ultra 9 285 CPU (24 cores) and 64 GB of RAM under Ubuntu 24.04.5 LTS. The solver was Newton 1.2.1 on NVIDIA Warp 1.13.0, built against CUDA 12.9. Isaac Sim 6.0.1 rendered the probe, leg and room, and the Isaac for Healthcare ray tracer (`i4h-sensor-simulation` 0.4.0) ran in a separate container on CUDA 12.8 [5].

C. Probe Contact and Measurements

Probe–tissue contact is a linear penalty spring between the rigid probe and a contact sphere of radius 0.25 × element size on each tissue node,

$$\mathcal{E}_{\text{probe}} = \tfrac{1}{2}k_e d^2, \qquad \mathbf{f}_n = k_e d\,\mathbf{n}, \tag{3}$$

where d is the sphere's penetration of the probe surface, $\mathbf{n}$ the surface normal. $k_e = 2000$ N/m is a numerical penalty parameter tuned by hand so that the probe face did not visibly pass through the skin. A node of mass $m = \rho\,\Delta x^3$ on this spring oscillates with period $T = 2\pi\sqrt{m/k_e}$, the contact spring's period: 0.258 ms for 1.5 mm elements and $\rho = 1000$ kg/m^3, between the tabulated densities of fat and muscle [32]. Opposite vessel walls closer than $r = 0.4\times$ element size are held apart by a quadratic penalty and, below $\tau = r/2$, a logarithmic barrier. For two wall points δ apart, $\mathcal{E}_{\text{wall}} = \frac{1}{2}k_e(r-\delta)^2$ for $\tau \le \delta < r$ and $\frac{1}{2}k_e\tau^2[1 - 2\ln(\delta/\tau)]$ for $\delta < \tau$, with energy and force continuous at τ (Section S1). The contact energy in Eq. (2) is then

$$\mathcal{E}_{\text{contact}} = \sum_j \mathcal{E}_{\text{probe}}(d_j) + \sum_{(a,b)} \mathcal{E}_{\text{wall}}(\delta_{ab}), \tag{4}$$

over tissue nodes j penetrating the probe and wall-point pairs (a, b) closer than r. The walls never meet, so the mesh cannot represent a lumen area below a minimum, the *closure floor* (Section III-D).

The probe, a flat 40×12 mm face, was displacement-controlled. A sweep first rotated the drawn probe from $-45°$ to $+45°$ about the limb axis at $s = 20$ mm, hovering 2 mm above the skin, so only the image changed (Fig. S3). The press then advanced at $0°$ by 0.4 mm per frame to 3, 7, 11, 15, 19, 23 and 25.4 mm, holding each depth until it settled. Lumen area is measured on the $s = 20$ mm cut (Section S1). Probe force can be derived from the contact term but is not reported (Section IV-D).

D. Benchtop Phantom: A Low-Fidelity Feasibility Test

The simulation was first tested on a benchtop phantom molded from the leg design (Section II-E, Table I). The phantom is deliberately low in fidelity, but its materials have supplier-specified properties. It has a muscle core of radius 76.2 mm inside a 15.9 mm fat layer, 81 mm long, both blends of a synthetic tissue gel [33] whose Young's moduli, 510 and 380 kPa, are the supplier's values mixed by mass. Silicone tubes stand in for the vessels: the artery tube has a 9 mm bore and a 1.0 mm wall of Durometer 70A, the vein tube a 10 mm bore and a 1.5 mm wall of 35A. Their moduli, 5,500 and




Fig. 1. Phantom validation setup. Top: the simulated phantom, cut open at the ultrasound plane, and its ray-traced B-mode image. Bottom: the UR5 robot pressing the linear probe on the benchtop phantom, and the scanner's B-mode image. Vein (blue) and artery (red) are labelled in both images.

TABLE I

THE BENCHTOP PHANTOM AGAINST THE ANATOMICAL MODEL. LEG MODEL (CAD): THE CROSS-SECTION 20 MM BELOW THE INGUINAL LIGAMENT, AS MESHED, WITH THE BENCHTOP PHANTOM'S TWO SIMPLIFICATIONS (NO PUBIC RAMUS; 81 MM LONG); MODULI IN TABLE II. BENCHTOP: THE MOLD'S DESIGN DIMENSIONS AND CATALOGUE TUBE SIZES, NOT CALIPER MEASUREMENTS OF THE BENCHTOP PHANTOM; MODULI ARE SUPPLIER VALUES. IT HAS NO SHEATH, FEMORAL CANAL OR NERVE.

Parameter	**Leg model (CAD)**	**Benchtop**
Length along the limb	81 mm	81 mm
Diameter	184.9 mm	184.1 mm
Skin and fat thickness	16.0 mm	15.9 mm
Artery lumen diameter	6.96 mm	9.00 mm
Vein lumen diameter	9.95 mm	10.00 mm
Artery wall thickness	1.00 mm	1.00 mm
Vein wall thickness	0.50 mm	1.50 mm
Skin to artery inner wall	19.5 mm	22.6 mm
Skin to vein inner wall	17.8 mm	20.8 mm
Artery to vein, center to center	10.0 mm	16.0 mm
Tissue model	Neo-Hookean; moduli in Table II	Neo-Hookean; values below
Soft-tissue material	–	SimuGel blends (muscle #0+#1, fat #2+#0)
Young's modulus, muscle / fat	–	510 / 380 kPa (supplier, mixed by mass)
Vessel walls	–	silicone tubes, 70A artery / 35A vein

1,400 kPa, follow from durometer by Gent's relation [34]. The tubes are water-filled and open at both ends, so no pressure builds. The femur is a printed copy held only by the gel, and there is no sheath, femoral canal or nerve. Its model, the simulated phantom (Fig. 1), used the leg's element sizes and solver settings (Section II-F), $\nu = 0.49$ for gels and tubes, as for nearly incompressible solids [35], [36], a free femur at 200,000 kPa, and for walls thinner than one element the modulus that keeps their bending stiffness, $E_{\text{eff}} = E\,(t/1.5\ \text{mm})^3$ (Table S6). No input was adjusted to the phantom.

A UR5 robot (Universal Robots, Odense, Denmark) held a linear ultrasound probe on the phantom 20 mm below the ligament's level. Zero was the first probe height that gave a visible image. The probe was pressed to the seven depths three times each, withdrawn between depths, and held 5 s before each image. Inner lumen height was measured between the inner wall echoes at one column per vessel: a Python script (NumPy, SciPy) found the echo peaks and the operator chose the inner pair (script, picks and every image in the supplement, Fig. S7). Height was converted with the scanner's 22.24 pixels/mm and the 1480/1540 m/s ratio of sound speeds in water and in the scanner's model, then to area through a rounded-rectangle model, $A/A_0 = x(2-x)$ with $x = h/D_0$ the inner height over the catalogue bore. Heights at first contact were 8 to 10% shorter and served as a sensitivity check. Agreement at a depth was set before the presses at 5 points, close agreement at 3, on the same basis as the numerical tolerance [37].

E. Leg Model: Anatomy and Materials

The anatomy is a parametric CAD model of the right femoral region, dimensioned from the literature (Table II). The segment runs from the inguinal ligament to $s = 81$ mm, the length of the benchtop phantom. Its reference level, $s = 20$ mm, lies inside the arterial window that ends at the profunda femoris origin [38]. It is a model level, not a recommended puncture site: a safe arterial puncture lies on the common femoral artery below the inguinal ligament and above its bifurcation, located on each patient's ultrasound [38], [39]. The mesh extrudes the $s = 20$ mm cross-section along the limb, so each body has a constant section (Table II). Vessel lumen targets are the duplex ultrasound means of Zambetti *et al.* ($n = 983$) [40], measured at the saphenofemoral junction, a different level from ours. Artery diameter also varies along the vessel [41] and with age, sex and body size [42]. As built, both lumens are smaller than their targets because the CAD read each target lumen diameter as the outer wall diameter: the common femoral artery (CFA) lumen is 7.0 mm against 9.0 ± 2.1 mm and the common femoral vein (CFV) lumen 10.0 mm against 11.0 ± 2.7 mm [40]. Skin-to-intima depths, 19.5 mm (CFA) and 17.8 mm (CFV), lie within 0.35 SD of the cohort means [40]. The femur is the only bone and its nodes are pinned. Every modulus and Poisson ratio is an informed modeling choice made with reference to published measurements of each tissue [35], [43]–[56] (Section IV-D). The vessels are treated

TABLE II

STRUCTURES MODELED, AS MESHED AT $s = 20$ MM. DEPTHS ARE PERPENDICULAR TO THE SKIN, EXCEPT THE CFV OUTER FAR WALL. ITS 29.2 MM IS VERTICAL, BELOW THE FRONT OF THE THIGH, AND 28.2 MM PERPENDICULAR TO THE SKIN. (SET): A DIMENSION ASSIGNED FOR THE MODEL WHERE NO MEASUREMENT APPLIED, AS IS THE 10.0 MM ARTERY–VEIN CENTER SPACING. EACH E AND ν WAS CHOSEN WITH REFERENCE TO PUBLISHED MEASUREMENTS (SECTION IV-D). E IS NOMINAL, AND ELEMENTS AT MATERIAL INTERFACES TAKE BLENDED VALUES.

Structure	**Geometry (source)**	E **(kPa)**	ν
Skin and fat (one body)	shell 16.0 mm; limb diameter 184.9 mm (set)	3	0.45
Muscle	core, radius 76.5 mm	20	0.45
CFA wall	lumen 7.0 mm, wall 1.0 mm (set); intima 19.5 mm deep [40]	600	0.45
CFV wall	lumen 10.0 mm, wall 0.5 mm (set); intima 17.8 mm deep [40]; outer far wall 29.2 mm vertical (28.2 mm perpendicular)	20	0.45
Femoral nerve	ellipse 9.6×3.7 mm [57]; 13.0 mm lateral to CFA (about 12.7 mm [58]; measured 11.1 ± 2.9 mm [59])	40	0.45
Femoral sheath	constant section in the mesh (funnel in CAD only); $s = 0$–35 mm (set; usually 30–40 mm long [38]); wall thickness set	20	0.45
Perivascular tissue	fills the sheath around the vessels	3	0.45
Femur	23.4 mm rod on the limb axis (set)	pinned	–

as voids with no internal pressure.

TABLE III

GEOMETRY AND NUMERICAL SETTINGS OF THE SLAB AND LEG MESHES. SIZES ARE THE BOUNDING BOX OF THE RESTING MESH, WITH LATERAL ACROSS THE LIMB, AXIAL ALONG IT AND DEPTH FROM THE SKIN INWARD. THE LEG COLUMN GIVES THE CHOSEN SETTING; THE REFERENCE SETTING USES A 0.099 MS TIME STEP (168 TIME STEPS PER FRAME) WITH 140 ITERATIONS PER STEP. (SET): ASSIGNED FOR THE MODEL. THE DAMPING COEFFICIENT γ IS IN SECONDS, AS IN THE SOLVER'S DAMPING FORM (SECTION S1).

Parameter	**Slab (parameter study)**	**Leg model**
Domain	block cut from the limb 20 mm below the inguinal ligament; side and deep faces held fixed	whole thigh segment, 81 mm long, held by a small fixed patch where the thigh rests on the table
Lateral width	72.0 mm	184.9 mm
Axial length	16.5 mm	81.0 mm
Depth	82.5 mm	184.9 mm
Nodes	31,740	23,688
Tetrahedra	166,386	129,783
Nodes held fixed	1,896 (side and deep faces, bone)	352 (bone 342, table patch 10)
Element size	1.5 mm, uniform	1.5 mm near the vessels; 5 mm femur; 30 mm elsewhere
Time step	0.0186 ms (896 per frame)	8.33 ms (2 per frame)
Iterations per time step	140	35
Contact stiffness k_e	2,000 N/m	2,000 N/m
Damping γ (MC)	1.0 s	1.0 s




Fig. 2. Benchtop phantom against the simulated phantom: lumen area against probe depth, at 20 mm below the inguinal ligament. Lines: simulated phantom, reference setting. Markers: benchtop phantom, mean ± SD of three robot presses (one 25.4 mm press from a separate session). Bands: ±5 points about each simulated line, the agreement tolerance set before the presses.

F. Solver Calibration

Numerical parameters were first studied on a slab cut from the leg at $s = 20$ mm (Table III), then on the leg. The slab is a rectangular cut of the leg's cross-section on a regular grid, with the artery, vein and nerve at the same positions within 0.55 mm (Fig. S3). Element sizes of 2.5, 2.0, 1.5, 1.25 and 1.0 mm were run at a 0.0186 ms time step with 140 iterations per step. Since settling shows only that a run has stopped moving, mesh dependence was judged by comparing vein closure between sizes. Accuracy was judged against a reference run on each mesh: on the slab the 0.0186 ms, 140-iteration run. On the leg, it is a 0.099 ms time step with 140 iterations, the finest leg setting that completed without GPU failure. Both are numerical references, not exact solutions, and both still change with element size. A setting passes when vein and artery lumen areas stay within 3 points of the reference at every depth. This tolerance was set before the repeat runs and follows the 2 to 5% bands that define settling time in controls theory [37]. The leg mesh is graded, with 1.5 mm elements near the vessels and coarser ones elsewhere (Table III). Each solver iteration is a fixed series of GPU kernel launches, which the CPU otherwise issues one by one. A CUDA Graph records the series once, and the CPU then submits it with a single call [60].

III. RESULTS

A. Benchtop Phantom

Fig. 2 compares the two phantoms (Table S5, with every image in Fig. S7). The artery agreed within 5 points at all seven depths, and within 3 at six of them. The vein

agreed within 3 points through 11 mm. From 15 to 25.4 mm, the simulated phantom collapsed the vein more than the benchtop phantom did, by 5.3 to 11.9 points, always with the same sign. With the first-contact heights as rest diameters, the artery still agreed within 5 points at every depth while the vein discrepancies grew to at most 16.4 points. Neither phantom separated the vessels. At every depth their vein and artery lumens stayed within 8 points of each other on the benchtop and 5 in simulation. On this material, about 25 times stiffer than soft tissue, the chosen setting differed from the reference by up to 4.4 points, so the comparison uses the reference setting.

B. Slab: Element Size, Time Step and Iterations

The slab set the element size, time step and iteration count carried to the leg (Section S2). At a 0.0186 ms time step, vein closure at 25.4 mm still changed with element size, from 33.8% of rest at 1.5 mm to 19.5% at 1.0 mm (Fig. S4, Table S1). The 1.5 mm size was chosen on cost, although a finer mesh still changed the result. Coarsening the time step 448-fold, to 8.33 ms, moved vein closure by 1.89 ± 0.02 points over three runs (Fig. S5, Table S2). Two settings failed to settle at some depths: their time steps, 0.249 and 0.980 ms, are 0.96 and 3.80 times the contact spring's 0.258 ms period. At 2.08 ms, 35 iterations stayed within 0.39 ± 0.05 points of the 140-iteration reference, so the leg model used 35 iterations per step.

C. Leg: Speed and Accuracy

CUDA Graphs removed the host's per-launch cost, a 3-fold gain at equal settings (Table S3), and moved closure at 25.4 mm by under 1 point. Fig. 3 and Table S4 give every leg setting. Wall clock followed solver work at 0.400 ± 0.001 ms per solver pass (Table S3). Accuracy did not follow the work increase. All twelve settings stayed inside the 3-point tolerance, and the largest gap, 1.20 points, belonged to the 2.08 ms step with 35 iterations, which did four times the chosen setting's work. The chosen setting, solver only, took 30.4 ms per frame, 977 times faster than the reference without CUDA Graphs. Over five runs it stayed within 0.76 ± 0.09 (vein) and 1.03 ± 0.00 (artery) points of the reference mean at every depth, while repeat runs of the reference differed from their mean by 0.23 ± 0.16 and 0.13 ± 0.06 points. The full loop, adding ray-traced ultrasound and the rendered 3-D scene, took 56.4 ms per frame in one run and stayed within 0.60 (vein) and 1.22 (artery) points of the reference mean. Every leg setting settled at every depth in all five runs.

D. Predicted Vessel Response in the Leg Model

The vein compressed more than the artery at every depth and every solver setting (Fig. 4). At 11 mm the vein kept 75.4% of its resting area and the artery 99.7% (five runs). At 25.4 mm they kept $27.7 \pm 0.1\%$ and $83.7 \pm 0.1\%$, 56 points apart, where neither phantom's vessels differed by more than 8. At 11 mm the vessels had already separated by 24 points while the vein kept three quarters of its area as a target for the needle. How much separation a robot needs




Fig. 3. Leg model on one RTX 5090 with no other GPU load: wall clock per simulated frame against solver work per frame (time steps per frame × iterations per step), with the equivalent frames per second on the right. Filled circles: the twelve solver settings with CUDA Graphs on, mean of five runs (SD 0.4–2.3% of the mean, smaller than the marker, Table S4). Two settings share work 280. Open circles: CUDA Graphs off (Table S3). "Initial setting": the reference setting. Filled star, "final run": the chosen setting, solver only (30.4 ms per frame). Open star: the chosen setting with the 3-D view and ray-traced ultrasound on the same GPU (56.4 ms, one run). Dashed: the 16.7 ms real-time limit. Line: two-term fit to the sixty CUDA-Graphs-on runs, 0.401 ms per solver pass plus 2.82 ms per frame (Fig. S6, Table S3).

is not established here. The curve is indexed by probe depth without force, in vessels that hold no blood pressure, so it cannot predict the depth that suits a given patient. The vein reading stayed above zero because contact holds opposite walls about 1.2 mm apart ($0.8\times$ element size). This sets a nominal closure floor of 14–20% on this mesh, and the reading stayed above it. On the leg, at most 859 elements (0.67% of the mesh) were inverted at 25.4 mm, none in a vessel wall (Section S2).

IV. Discussion

A. Agreement with the Benchtop Phantom

The benchtop test asked whether the simulation could reproduce a simple object of known materials without tuning. It reproduced the artery at every depth and the vein through 11 mm. From 15 mm on the simulated vein always collapsed more than the benchtop vein. The discrepancy most likely comes from the vein tube's wall stiffness relative to the gel (Section S3).

B. From Phantom to Clinical Parameters

Neither phantom separated the vessels, so this test checked the solver, not the clinical signal. Once the solver had passed it, we gave the leg model literature-based tissue properties rather than building a more faithful phantom. The leg then




Fig. 4. Leg model: vein (blue) and artery (red) lumen area against probe depth, for all twelve solver settings searched. Thin lines: the ten intermediate settings. Purple dashed: the reference setting. Green stars: the chosen setting (± SD). Each is the mean of five runs. Hollow green diamonds: one run of the full loop. Gray band: the closure floor, the nominal reading of a fully flattened vein on this mesh, derived from contact spacing. No run reached it, and the vein compressed but never closed.

compressed the vein far more than the artery at every solver setting (Section III-D). The clinical signal came from the clinical inputs, not from the solver's cost. Whether the leg model carries the phantom's vein bias is untested, because its materials differ. Its predicted vein collapse at depth is a property of the model, not yet of a patient.

C. Speed and Accuracy Trade-off

On the leg, an 84-fold coarser time step and a quarter of the iterations kept closure well inside the tolerance set before the runs, above the reference's run-to-run spread. This is agreement with a numerical reference on one mesh, not with real tissue. On the slab, element size mattered more than any settled time step (Section III-B). We therefore suggest choosing the cheapest setting that settles at every depth and stays within tolerance of the reference, repeating both checks for every new mesh, material or probe motion. At 56.4 ms per frame, the full loop is slower than the 16.7 ms real-time limit but within the 43 to 100 ms between actions of diffusion-policy controllers for robotic ultrasound [61]. The reference, at 29,700 ms per frame, falls far outside that range.

D. Limitations

The vessels hold no blood pressure, flow, nor pulsation, which under ultrasound marks the artery, though a nearby vein can appear to pulse with it. In canines, venous pressure widens the femoral vein [62], rounding its section [63]. Arterial pressure would resist compression far more, so we expect pressure to widen the difference between the vessels. This omission is most serious in the emergencies we target. In shock, vein size changes with fluid status. In cardiac arrest the artery loses its pulse, and below a systolic pressure of 60 mmHg it can also compress under the probe [18]. Performance in these emergency states remains unknown.

No published stiffness was measured under this loading (Section S3, Table S6). Compared by shear modulus ($\nu = 0.45$), our fat (1.0 kPa) is about twice the in vivo thigh median and our muscle (6.9 kPa) about quadruple [43]. The vein wall departs furthest from published data. Tensile tests give 5,700 kPa [51], about 285 times our 20 kPa, which we chose to represent a thin wall's collapse under low transmural pressure [63]. Bending stiffness scales with Et^3, and our one-element wall is 1.5 mm thick against 0.33–0.35 mm in vivo [64]. Walls matching the bending stiffness of a 0.34 or 0.50 mm wall (66.4 or 211.1 kPa) kept 35.3% and 51.4% of the vein's area at 25.4 mm against 27.6% for 20 kPa, with the artery at 83.6–86.2% in all three (Table S6). Thus, the vein closes later but still compresses far more than the artery.

We do not report probe force. The endpoint is read from the image, the force error grows as elements invert (Section S2), and force was never calibrated against a known load. The benchtop comparison rests on one phantom at one level, with unmeasured supplier and durometer moduli and areas inferred from one inner height. Gel and silicone agreement does not show that the model predicts human tissue. The acoustic materials are NVIDIA's seven liver-workflow defaults [5], and the simulated image has not been compared with clinical images.

V. Conclusion

A simulation in Isaac Sim predicted a low-fidelity benchtop phantom's artery within 5 points at every depth. It matched the vein within 3 points through 11 mm, then collapsed the vein more than the phantom did. Robotic interaction does not require the finest solver setting, only an error small enough for the task. The solver's time per frame fell from 29,700 ms to 30.4 ms while the chosen setting stayed within 1.03 points of the reference. Because accuracy did not improve with more work, each candidate setting must be checked against a reference. Given literature-based tissue properties, the leg model kept the differential vein and artery deformation the ultrasound observation needs at every setting. Future work will add blood pressure and flow, validated force measurement and acoustic validation, then a physical robot in the loop on the benchtop phantom. With needle insertion added, success will mean first-attempt entry into the intended vessel [39], no back-wall puncture and a guidewire inside the lumen [18].

Acknowledgment

The authors thank Professor Karthik Desingh and the Robotics: Perception and Manipulation Laboratory for computing resources.

REFERENCES

[1] Z. Jiang, S. E. Salcudean, and N. Navab, "Robotic ultrasound imaging: state-of-the-art and future perspectives," *Med. Image Anal.*, vol. 89, p. 102878, 2023.

[2] H. Elmekki, S. Islam, A. Alagha, H. Sami, A. Spilkin, E. Zakeri, A. M. Zanuttini, J. Bentahar, L. Kadem, W.-F. Xie, P. Pibarot, R. Mizouni, H. Otrok, S. Singh, and A. Mourad, "Comprehensive review of reinforcement learning for medical ultrasound imaging," *Artif. Intell. Rev.*, vol. 58, no. 9, p. 284, 2025.

[3] D. Guo, Y. Zhang, X. Huang, and C. He, "Technical challenges in autonomous robotic ultrasound examinations: perception, planning, and control," *J. Robot. Surg.*, vol. 20, no. 1, 2026.

[4] M. Mittal, C. Yu, Q. Yu, J. Liu, N. Rudin, D. Hoeller, J. L. Yuan, R. Singh, Y. Guo, H. Mazhar, A. Mandlekar, B. Babich, G. State, M. Hutter, and A. Garg, "Orbit: a unified simulation framework for interactive robot learning environments," *IEEE Robot. Autom. Lett.*, vol. 8, no. 6, pp. 3740–3747, 2023.

[5] NVIDIA Corporation, "Isaac Sim and Isaac for Healthcare (documentation and code)," https://github.com/isaac-for-healthcare/i4h-workflows, 2026.

[6] E. Aljalbout, J. Xing, A. Romero, I. Akinola, C. R. Garrett, E. Heiden, A. Gupta, T. Hermans, Y. Narang, D. Fox, D. Scaramuzza, and F. Ramos, "The reality gap in robotics: challenges, solutions, and best practices," *Annu. Rev. Control Robot. Auton. Syst.*, vol. 9, pp. 403–432, 2026.

[7] A. Zafra-Navarro, R. Aragüés, and G. López-Nicolás, "Survey of simulators for deformable objects in robotics," in *Jornadas de Automática*, vol. 46, 2025.

[8] C. Shen, Z. Deng *et al.*, "Towards autonomous robotic ultrasound scanning using the reinforcement learning-based volumetric data navigation method," in *Proc. WRC Symp. Adv. Robot. Autom. (WRC SARA)*, 2023, pp. 335–340.

[9] C. Luo, Y. Chen *et al.*, "Multi-modal autonomous ultrasound scanning for efficient human-machine fusion interaction," *IEEE Trans. Autom. Sci. Eng.*, vol. 22, pp. 4712–4723, 2025.

[10] Y. Ou and M. Tavakoli, "CRESSim-MPM: a material point method library for surgical soft body simulation with cutting and suturing," in *Proc. IEEE/RSJ Int. Conf. Intell. Robots Syst. (IROS)*, 2025, arXiv:2502.18437.

[11] X. Liang, F. Liu, Y. Zhang, Y. Li, S. Lin, and M. Yip, "Real-to-sim deformable object manipulation: optimizing physics models with residual mappings for robotic surgery," in *Proc. IEEE Int. Conf. Robot. Autom. (ICRA)*, 2024, arXiv:2309.11656.

[12] Y. Qin, L. Dang *et al.*, "A review and perspective of techniques for autonomous robotic ultrasound acquisitions," *Sensors*, vol. 26, no. 7, p. 2081, 2026.

[13] Z. Jiang, Y. Gao, L. Xie, and N. Navab, "Towards autonomous atlas-based ultrasound acquisitions in presence of articulated motion," *IEEE Robot. Autom. Lett.*, vol. 7, no. 3, pp. 7423–7430, 2022.

[14] F. Šuligoj, C. M. Heunis, J. Sikorski, and S. Misra, "RobUSt: an autonomous robotic ultrasound system for medical imaging," *IEEE Access*, vol. 9, pp. 67 456–67 465, 2021.

[15] Y. Qian, J. Luo, W. Zhu, Y. Zhang, M. Q.-H. Meng, Y. Yuan, and L. Liu, "Task-relevant feature-dynamics fidelity enables zero-shot sim-to-real transfer for robotic ultrasound scanning," arXiv:2608.29516, 2026.

[16] A. Stannard, J. L. Eliason, and T. E. Rasmussen, "Resuscitative endovascular balloon occlusion of the aorta (REBOA) as an adjunct for hemorrhagic shock," *J. Trauma*, vol. 71, no. 6, pp. 1869–1872, 2011.

[17] J. Swol, J. Belohlávek, D. Brodie *et al.*, "Extracorporeal life support in the emergency department: a narrative review for the emergency physician," *Resuscitation*, vol. 133, pp. 108–117, 2018.

[18] B. Saugel, T. W. L. Scheeren, and J.-L. Teboul, "Ultrasound-guided central venous catheter placement: a structured review and recommendations for clinical practice," *Crit. Care*, vol. 21, no. 1, p. 225, 2017.

[19] L. J. Brattain, T. T. Pierce *et al.*, "AI-enabled, ultrasound-guided handheld robotic device for femoral vascular access," *Biosensors*, vol. 11, no. 12, p. 522, 2021.

[20] J. Guerrero, S. E. Salcudean, J. A. McEwen, B. A. Masri, and S. Nicolaou, "System for deep venous thrombosis detection using objective compression measures," *IEEE Trans. Biomed. Eng.*, vol. 53, no. 5, pp. 845–854, 2006.

[21] M. Lamperti, A. R. Bodenham *et al.*, "International evidence-based recommendations on ultrasound-guided vascular access," *Intensive Care Med.*, vol. 38, no. 7, pp. 1105–1117, 2012.

[22] A. Vegas, B. Wells, P. Braum, A. Denault, W. C. Miller Hance, C. Kaufman, M. B. Patel, and M. Salvatori, "Guidelines for performing ultrasound-guided vascular cannulation: recommendations of the American Society of Echocardiography," *J. Am. Soc. Echocardiogr.*, vol. 38, no. 2, pp. 57–91, 2025.

[23] X. Bao, S. Wang *et al.*, "A constant-force end-effector with online force adjustment for robotic ultrasonography," *IEEE Robot. Autom. Lett.*, vol. 6, no. 2, pp. 2547–2554, 2021.

[24] L. Beber, E. Lamon, D. Nardi, D. Fontanelli, M. Saveriano, and L. Palopoli, "A passive variable impedance control strategy with viscoelastic parameters estimation of soft tissues for safe ultrasonography," in *Proc. IEEE Int. Conf. Robot. Autom. (ICRA)*, 2024, pp. 1298–1304, arXiv:2309.14893.

[25] W. Si, N. Wang, and C. Yang, "Design and quantitative assessment of teleoperation-based human-robot collaboration method for robot-assisted sonography," *IEEE Trans. Autom. Sci. Eng.*, vol. 22, pp. 317–327, 2025.

[26] D. Raina, L. Al-Zogbi *et al.*, "AURA-CVC: autonomous ultrasound-guided robotic assistance for central venous catheterization," *Int. J. Comput. Assist. Radiol. Surg.*, 2026, online first.

[27] F. Alvarado, "How to do femoral vein cannulation, ultrasound-guided," MSD Manual Professional Version, Merck & Co., Rahway, NJ, USA, 2026, https://www.merckmanuals.com/professional/critical-care-medicine/how-to-do-central-vascular-procedures/how-to-do-femoral-vein-cannulation-ultrasound-guided, accessed Nov. 2, 2026.

[28] S. I. Seldinger, "Catheter replacement of the needle in percutaneous arteriography; a new technique," *Acta Radiol.*, vol. 39, no. 5, pp. 368–376, 1953.

[29] B. Smith, F. de Goes, and T. Kim, "Stable Neo-Hookean flesh simulation," *ACM Trans. Graph.*, vol. 37, no. 2, pp. 12:1–12:15, 2018.

[30] A. H. Chen, Z. Liu, Y. Yang, and C. Yuksel, "Vertex block descent," *ACM Trans. Graph.*, vol. 43, no. 4, pp. 116:1–116:16, 2024.

[31] G. Irving, J. Teran, and R. Fedkiw, "Tetrahedral and hexahedral invertible finite elements," *Graph. Models*, vol. 68, no. 2, pp. 66–89, 2006.

[32] C. Baumgartner, P. A. Hasgall, F. Di Gennaro *et al.*, "IT'IS database for thermal and electromagnetic parameters of biological tissues, version 5.0," IT'IS Foundation, Zurich, 2025, accessed 2026-10-28.

[33] Humimic Medical, "SimuGel technical documentation, version 2.3," Greenville, SC, USA, 2026.

[34] A. N. Gent, "On the relation between indentation hardness and Young's modulus," *Rubber Chem. Technol.*, vol. 31, no. 4, pp. 896–906, 1958.

[35] P. H. Mott, J. R. Dorgan, and C. M. Roland, "The bulk modulus and Poisson's ratio of "incompressible" materials," *J. Sound Vib.*, vol. 312, no. 4–5, pp. 572–575, 2008.

[36] N. Fougeron, A. Trebbi, B. Keenan, Y. Payan, and G. Chagnon, "Current Poisson's ratio values of finite element models are too low to consider soft tissues nearly-incompressible: illustration on the human heel region," *Comput. Methods Biomech. Biomed. Eng.*, vol. 27, no. 14, pp. 1999–2008, 2024.

[37] K. Ogata, *Modern Control Engineering*, 5th ed. Upper Saddle River, NJ: Prentice Hall, 2010, settling time defined by a 2% or 5% band about the final value.

[38] K. L. Moore, A. F. Dalley, II, and A. M. R. Agur, *Clinically Oriented Anatomy*, 8th ed. Philadelphia, PA: Wolters Kluwer, 2018, excerpt on the femoral triangle and femoral sheath.

[39] A. H. Seto, M. S. Abu-Fadel *et al.*, "Real-time ultrasound guidance facilitates femoral arterial access and reduces vascular complications: FAUST (Femoral Arterial Access With Ultrasound Trial)," *JACC Cardiovasc. Interv.*, vol. 3, no. 7, pp. 751–758, 2010.

[40] B. R. Zambetti, A. Kankaria, F. Fang, N. Kim, K. Nagarsheth, and R. Sarkar, "Diameter and depth of femoral vessels by duplex ultrasound," *J. Vasc. Access*, vol. 26, no. 1, pp. 162–167, 2025.

[41] I. Prună (Ion), C. Ciufu, and P. Bordei, "Common femoral artery diameters determined by Doppler ultrasonography," *ARS Medica Tomitana*, vol. 25, no. 1, pp. 1–5, 2019.

[42] T. Sandgren, B. Sonesson, Å. Rydén Ahlgren, and T. Länne, "The diameter of the common femoral artery in healthy human: influence of sex, age, and body size," *J. Vasc. Surg.*, vol. 29, no. 3, pp. 503–510, 1999.

[43] A. Segain, G. Sciume, H. Pillet, and P.-Y. Rohan, "In vivo mechanical response of thigh soft tissues under compression: A two-layer model allows an improved representation of the local tissue kinematics," *Journal of the Mechanical Behavior of Biomedical Materials*, vol. 156, p. 106584, 2024.

[44] M. K. Chakouch, F. Charleux, and S. F. Bensamoun, "Quantifying the elastic property of nine thigh muscles using magnetic resonance elastography," *PLoS ONE*, vol. 10, no. 9, p. e0138873, 2015.

[45] S. F. Bensamoun, S. I. Ringleb, L. Littrell, Q. Chen, M. Brennan, R. L. Ehman, and K.-N. An, "Determination of thigh muscle stiffness using magnetic resonance elastography," *J. Magn. Reson. Imaging*, vol. 23, no. 2, pp. 242–247, 2006.

[46] N. Alkhouli, J. Mansfield, E. Green, J. Bell, B. Knight, N. Liversedge, J. C. Tham, R. Welbourn, A. C. Shore, K. Kos, and C. P. Winlove, "The mechanical properties of human adipose tissues and their relationships to the structure and composition of the extracellular matrix," *Am. J. Physiol. Endocrinol. Metab.*, vol. 305, no. 12, pp. E1427–E1435, 2013.

[47] A. Samani, J. Zubovits, and D. Plewes, "Elastic moduli of normal and pathological human breast tissues: an inversion-technique-based investigation of 169 samples," *Phys. Med. Biol.*, vol. 52, no. 6, pp. 1565–1576, 2007.

[48] X. Feng, G.-Y. Li, A. Ramier, A. M. Eltony, and S.-H. Yun, "In vivo stiffness measurement of epidermis, dermis, and hypodermis using broadband rayleigh-wave optical coherence elastography," *Acta Biomaterialia*, vol. 146, pp. 295–305, 2022.

[49] M. Jadidi, A. Desyatova, J. MacTaggart, and A. Kamenskiy, "Mechanical stresses associated with flattening of human femoropopliteal artery specimens during planar biaxial testing and their effects on the calculated physiologic stress-stretch state," *Biomechanics and Modeling in Mechanobiology*, vol. 18, no. 6, pp. 1591–1605, 2019.

[50] C. A. J. Schulze-Bauer, P. Regitnig, and G. A. Holzapfel, "Mechanics of the human femoral adventitia including the high-pressure response," *Am. J. Physiol. Heart Circ. Physiol.*, vol. 282, pp. H2427–H2440, 2002.

[51] K. Lebahn, J. Keiler, W. Schmidt, J. Schubert, M. Reumann, A. Wree, N. Grabow, and S. Kischkel, "Mechanical characterization of the human femoral vein wall and its valves," *Journal of the Mechanical Behavior of Biomedical Materials*, vol. 165, p. 106938, 2025.

[52] R. L. R. Wesly, R. N. Vaishnav, J. C. A. Fuchs, D. J. Patel, and J. C. Greenfield, Jr., "Static linear and nonlinear elastic properties of normal and arterialized venous tissue in dog and man," *Circ. Res.*, vol. 37, no. 4, pp. 509–520, 1975.

[53] J. Veselý, L. Horný, H. Chlup, T. Adámek, M. Krajíček, and R. Žitný, "Constitutive modeling of human saphenous veins at overloading pressures," *J. Mech. Behav. Biomed. Mater.*, vol. 45, pp. 101–108, 2015.

[54] Y. Song, Y. Zhang, Y. Zhang, and B. Hu, "Two-dimensional ultrasound and two-dimensional shear wave elastography on femoral and saphenous neuropathy in patients with type 2 diabetes mellitus," *Frontiers in Neurology*, vol. 13, p. 996199, 2022.

[55] O. Sánchez-Ramirez, Á. González-de-la Flor, C. Cotteret, and J. Almazán-Polo, "Within-session repeatability of femoral nerve shear wave elastography across fixed sequential Thomas test-derived protocol in healthy young adults," *J. Back Musculoskelet. Rehabil.*, 2026, online first.

[56] S. Otsuka, T. Yakura, Y. Ohmichi, M. Ohmichi, M. Naito, T. Nakano, and Y. Kawakami, "Site specificity of mechanical and structural properties of human fascia lata and their gender differences: A cadaveric study," *Journal of Biomechanics*, vol. 77, pp. 69–75, 2018.

[57] T. K. Lonchena, K. McFadden, and S. L. Orebaugh, "Correlation of ultrasound appearance, gross anatomy, and histology of the femoral nerve at the femoral triangle," *Surg. Radiol. Anat.*, vol. 38, pp. 115–122, 2016.

[58] H. Gray, *Anatomy, Descriptive and Surgical*, T. P. Pick and R. Howden, Eds. Philadelphia, PA: Lea Brothers, 1905, new American ed., rev. by J. C. Da Costa.

[59] W. T. Muhly and S. L. Orebaugh, "Ultrasound evaluation of the anatomy of the vessels in relation to the femoral nerve at the femoral crease," *Surg. Radiol. Anat.*, vol. 33, pp. 491–494, 2011.

[60] NVIDIA Corporation, "CUDA C++ programming guide, section "CUDA graphs"," https://docs.nvidia.com/cuda/cuda-c-programming-guide/, 2026, accessed Nov. 3, 2026.

[61] Y. Qian, Y. Zhu, J. Luo, L. Liu, Y. Yuan, G. Ning, and H. Liao, "Breaking the latency barrier: Synergistic perception and control for high-frequency 3D ultrasound servoing," arXiv preprint arXiv:2511.00983, 2025, preprint; full text checked 2026-10-09.

[62] T. W. Morris, P. H. Abbrecht, and S. D. Leverett, Jr., "Diameter-pressure relationships in the unexposed femoral vein," *Am. J. Physiol.*, vol. 227, no. 4, pp. 782–788, 1974.

[63] J. Gero and M. Gerová, "Venous geometry as determined by venous pressure," *Experientia*, vol. 24, pp. 1021–1022, 1968.

[64] F. Alibaz-Oner, R. Ergelen, I. Ergenc, G. Seven, A. Yazıcı, A. Cefle, C. Bes, O. Atug, and H. Direskeneli, "Femoral vein wall thickness measurement may be a distinctive diagnostic tool to differentiate Behçet's disease with intestinal involvement and Crohn's disease," *Digestive Diseases and Sciences*, vol. 66, pp. 2750–2755, 2021.