

Pivotal Study Protocol Synopsis

Draft U.S. pivotal-study design for Solaris Arc v2.3

Company: Solaris Endovascular, Inc.	Prepared: 2026-05-27
Status: Synthetic clinical protocol summary	Demo Use: Data-room ingestion and citation testing

Study Design

- Prospective, multicenter, single-arm pivotal study.
- Planned enrollment: 140 subjects.
- Planned U.S. clinical sites: 12.
- Target first subject in: Q4 2026, dependent on site activation and regulatory alignment.

Primary Endpoint

- Technical success without device-related major adverse event through 30 days.
- Endpoint adjudication will be performed by an independent clinical events committee.

Secondary Endpoints

- Navigation time.
- Successful agent delivery.
- Device deficiency rate.
- Physician usability score.
- Rate of minor and major protocol deviations.

Operational Risks

- Site activation may be delayed by budget negotiations or IRB review timing.
- Training consistency must be monitored across operators.
- Clinical supply must be released only after manufacturing and quality controls are complete.

Demo data room packet. Company-specific Solaris information is synthetic unless explicitly identified as public-source context.