

Laboratory Report

Printed by: LabTech

bioMérieux Customer:

System #:

Isolate: 1709-1 (Approved)

Card Type: GN Bar Code: 2413030403027760 Testing Instrument: 000019E1EF45 (VK2C 18325)

Card Type: AST-GN97 Bar Code: 6873220503721658 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Bionumber: 0003451003500252

Organism Quantity:

Selected Organism: Pseudomonas aeruginosa

Comments:	

McFarland: 0.62 (0.50 - 0.63)

Identification Information	Card:	GN	Lot Number:	2413030403	Expires:	Mar 14, 2026 12:00 ICT
	Status:	Final	Analysis Time:	6.78 hours	Completed:	Dec 1, 2025 18:21 ICT
Organism Origin	VITEK 2					
Selected Organism	99% Probability		Pseudomonas aeruginosa			
	Bionumber: 0003451003500252		Confidence: Excellent identification			
Analysis Organisms and Tests to Separate:						
Analysis Messages: The following antibiotic(s) are not claimed: ESBL, Neomycin, Pradofloxacin,						
Contraindicating Typical Biopattern(s)						

Card is for veterinary use only McFarland: 0.62 (0.50 - 0.63)

Susceptibility Information	Card:	AST-GN97	Lot Number:	6873220503	Expires:	Sep 20, 2026 12:00 ICT
	Status:	Final	Analysis Time:	13.88 hours	Completed:	Dec 2, 2025 01:27 ICT
Antimicrobial	MIC	Interpretation	Antimicrobial	MIC	Interpretation	
ESBL			Gentamicin	<= 1	S	
Ampicillin			Neomycin			
Amoxicillin/Clavulanic Acid			Enrofloxacin	0.5	S	
Cefalexin			Marbofloxacin	<= 0.5	S	
Cefalotin			Pradofloxacin			
Cefpodoxime			Doxycycline			
Cefovecin			Tetracycline			
Ceftiofur			Nitrofurantoin			
Imipenem	2	S	Chloramphenicol			
Amikacin	<= 2	S	Trimethoprim/ Sulfamethoxazole			

AES Findings:	Last Modified: May 17, 2024 10:18 ICT Parameter Set: Global CLSI-based+Natural Resistance V2 17 NOV2017
Confidence Level:	Consistent

Installed VITEK 2 Systems Version: 9.04.4

MIC Interpretation Guideline: Global CLSI-based Nov 16th V2

AES Parameter Set Name: Global CLSI-based+Natural Resistance

V2 17 NOV2017

Therapeutic Interpretation Guideline: NATURAL RESISTANCE

AES Parameter Last Modified: May 17, 2024 10:18 ICT

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1709-1 (Approved)

Card Type: GN Bar Code: 2413030403027760 Testing Instrument: 000019E1EF45 (VK2C 18325)

Card Type: AST-GN97 Bar Code: 6873220503721658 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Bionumber: 0003451003500252

Organism Quantity:

Selected Organism: Pseudomonas aeruginosa

Action	Name (User ID)	Date/Time	Comment
--------	----------------	-----------	---------