

Laboratory Report

Printed by: LabTech

bioMérieux Customer:

System #:

Isolate: 1739-1 (Approved)

Card Type: GP Bar Code: 2423225603014229 Testing Instrument: 000019E1EF45 (VK2C 18325)

Card Type: AST-GP81 Bar Code: 3813162403170104 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Bionumber: 070402077763231

Organism Quantity:

Selected Organism: Staphylococcus pseudintermedius

Comments:	

McFarland: 0.50 (0.50 - 0.63)

Identification Information	Card:	GP	Lot Number:	2423225603	Expires:	Sep 25, 2026 12:00 ICT
	Status:	Final	Analysis Time:	7.98 hours	Completed:	Dec 6, 2025 23:33 ICT
Organism Origin	VITEK 2					
Selected Organism	Staphylococcus pseudintermedius					
	Bionumber: 070402077763231			Confidence: Low discrimination		
Analysis Organisms and Tests to Separate:						
Low Discrimination Organism						
Staphylococcus aureus VP(99),Pyrro.Ary.(1),						
Staphylococcus pseudintermedius VP(10),Pyrro.Ary.(99),						
Analysis Messages:						
The following antibiotic(s) are not claimed:						
Cefoxitin Screen,						
Contraindicating Typical Biopattern(s)						
Staphylococcus aureus URE(2),						
Staphylococcus pseudintermedius dMAN(1),						

Card is for veterinary use only McFarland: 0.50 (0.50 - 0.63)

Susceptibility Information	Card: AST-GP81	Lot Number: 3813162403	Expires: Jul 24, 2026 12:00 ICT
	Status: Final	Analysis Time: 10.18 hours	Completed: Dec 7, 2025 01:45 ICT

Antimicrobial	MIC	Interpretation	Antimicrobial	MIC	Interpretation
Cefoxitin Screen			Pradofloxacin	1	I
Benzylpenicillin	>= 0.5	R	Inducible Clindamycin Resistance	NEG	-
Amoxicillin/Clavulanic Acid	16	R	Erythromycin	>= 8	R
Oxacillin	>= 4	R	Clindamycin	>= 4	R
Cefalotin	>= 32	R	Doxycycline	8	R
Cefpodoxime	>= 8	R	Minocycline	4	R
Cefovecin	>= 8	R	Nitrofurantoin	<= 16	S
Amikacin	<= 2	S	Chloramphenicol	8	S
Gentamicin	8	I	Florfenicol	<= 4	S
Enrofloxacin	>= 4	R	Trimethoprim/Sulfamethoxazole	>= 320	R
Marbofloxacin	>= 4	R			

Installed VITEK 2 Systems Version: 9.04.4

MIC Interpretation Guideline: Global CLSI-based Nov 16th V2
 AES Parameter Set Name: Global CLSI-based+Natural Resistance
 V2 17 NOV2017

Therapeutic Interpretation Guideline: NATURAL RESISTANCE
 AES Parameter Last Modified: May 17, 2024 10:18 ICT

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1739-1 (Approved)

Card Type: GP Bar Code: 2423225603014229 Testing Instrument: 000019E1EF45 (VK2C 18325)

Card Type: AST-GP81 Bar Code: 3813162403170104 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Bionumber: 070402077763231

Organism Quantity:

Selected Organism: Staphylococcus pseudintermedius

AES Findings:	Last Modified: May 17, 2024 10:18 ICT Parameter Set: Global CLSI-based+Natural Resistance V2 17 NOV2017
Confidence Level:	Consistent

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1739-1 (Approved)

Card Type: GP Bar Code: 2423225603014229 Testing Instrument: 000019E1EF45 (VK2C 18325)

Card Type: AST-GP81 Bar Code: 3813162403170104 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Bionumber: 070402077763231

Organism Quantity:

Selected Organism: Staphylococcus pseudintermedius

Action	Name (User ID)	Date/Time	Comment
--------	----------------	-----------	---------