

Laboratory Report

Printed by: LabTech

bioMérieux Customer:

System #:

Isolate: 1772-1 (Approved)

Card Type: AST-GP81 Bar Code: 3813162403170100 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity: **Selected Organism: Staphylococcus pseudintermedius**

Comments:	

McFarland: 0.62 (0.50 - 0.63)

Identification Information	
Organism Origin	Technologist
Selected Organism	Staphylococcus pseudintermedius
	Entered: Dec 16, 2025 11:01 ICT By: LabTech
Analysis Messages: The following antibiotic(s) are not claimed: Cefoxitin Screen, This isolate is presumed to be resistant based on detection of inducible clindamycin resistance.	

Card is for veterinary use only McFarland: 0.62 (0.50 - 0.63)

Susceptibility Information	Card: AST-GP81	Lot Number: 3813162403	Expires: Jul 24, 2026 12:00 ICT
	Status: Final	Analysis Time: 9.50 hours	Completed: Dec 16, 2025 20:34 ICT

Antimicrobial	MIC	Interpretation	Antimicrobial	MIC	Interpretation
Cefoxitin Screen			Pradofloxacin	>= 4	R
Benzylpenicillin	>= 0.5	R	Inducible Clindamycin Resistance	POS	+
Amoxicillin/Clavulanic Acid	<= 2	*R	Erythromycin	>= 8	R
Oxacillin	>= 4	R	Clindamycin	0.25	*R
Cefalotin	<= 2	*R	Doxycycline	>= 16	R
Cefpodoxime	4	*R	Minocycline	2	R
Cefovecin	2	*R	Nitrofurantoin	<= 16	S
Amikacin	<= 2	S	Chloramphenicol	>= 64	R
Gentamicin	4	S	Florfenicol	<= 4	S
Enrofloxacin	>= 4	R	Trimethoprim/ Sulfamethoxazole	>= 320	R
Marbofloxacin	>= 4	R			

*= AES modified **= User modified

AES Findings:	Last Modified: May 17, 2024 10:18 ICT Parameter Set: Global CLSI-based+Natural Resistance V2 17 NOV2017
Confidence Level:	Consistent

Installed VITEK 2 Systems Version: 9.04.4

MIC Interpretation Guideline: Global CLSI-based Nov 16th V2

AES Parameter Set Name: Global CLSI-based+Natural Resistance

V2 17 NOV2017

Therapeutic Interpretation Guideline: NATURAL RESISTANCE

AES Parameter Last Modified: May 17, 2024 10:18 ICT

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1772-1 (Approved)

Card Type: AST-GP81 Bar Code: 3813162403170100 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity:

Selected Organism: Staphylococcus pseudintermedius

Action	Name (User ID)	Date/Time	Comment
Reviewed by:	(LabTech)	Dec 17, 2025 08:38 ICT	