

Laboratory Report

Printed by: LabTech

bioMérieux Customer:

System #:

Isolate: 1782-1 (Approved)

Card Type: AST-ST03 Bar Code: 5423289503201299 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity: **Selected Organism: Streptococcus canis**

Comments:	

Mcfarland: 0.52 (0.50 - 0.63)

Identification Information	
Organism Origin	Technologist
Selected Organism	Streptococcus canis
	Entered: Dec 16, 2025 14:57 ICT By: LabTech
Analysis Messages: The following antibiotic(s) are not claimed: Inducible Clindamycin Resistance, High level resistance to gentamicin (MIC of > 128 mg/L), is generally caused by the production of a bifunctional APH(2')-AAC(6) enzyme that determines loss of synergism of all aminoglycosides (except streptomycin and arbekacin) with b-lactams and glycopeptides irrespective of MIC values.	

Mcfarland: 0.52 (0.50 - 0.63)

Susceptibility Information	Card: AST-ST03	Lot Number: 5423289503	Expires: Nov 28, 2026 12:00 ICT
	Status: Final	Analysis Time: 20.50 hours	Completed: Dec 17, 2025 07:34 ICT

Antimicrobial	MIC	Interpretation	Antimicrobial	MIC	Interpretation
Benzylpenicillin	<= 0.06	S	Clindamycin	<= 0.25	S
Ampicillin	<= 0.25	S	Linezolid	<= 2	S
Cefotaxime	<= 0.12	S	Teicoplanin		
Ceftriaxone	<= 0.12	S	Vancomycin	<= 0.12	S
Gentamicin			Tetracycline	2	S
Levofloxacin	<= 0.25	S	Tigecycline	0.12	S
Moxifloxacin	0.12	S	Chloramphenicol	2	S
Inducible Clindamycin Resistance			Rifampicin		
Erythromycin	<= 0.12	S	Trimethoprim/ Sulfamethoxazole	<= 10	S

AES Findings:	Last Modified: May 17, 2024 10:18 ICT Parameter Set: Global CLSI-based+Natural Resistance V2 17 NOV2017
Confidence Level:	Consistent

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1782-1 (Approved)

Card Type: AST-ST03 Bar Code: 5423289503201299 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity:

Selected Organism: Streptococcus canis

Action	Name (User ID)	Date/Time	Comment
--------	----------------	-----------	---------