

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1814-1 (Approved)

Card Type: AST-GP81 Bar Code: 3813162403170376 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity:

Selected Organism: Staphylococcus pseudintermedius

Comments:	

McFarland: 0.50 (0.50 - 0.63)

Identification Information	
Organism Origin	Technologist
Selected Organism	Staphylococcus pseudintermedius
	Entered: Dec 22, 2025 16:16 ICT By: LabTech
Analysis Messages:	
The following antibiotic(s) are not claimed: Cefoxitin Screen,	

Card is for veterinary use only McFarland: 0.50 (0.50 - 0.63)

Susceptibility Information	Card: AST-GP81	Lot Number: 3813162403	Expires: Jul 24, 2026 12:00 ICT
	Status: Final	Analysis Time: 14.33 hours	Completed: Dec 23, 2025 06:39 ICT

Antimicrobial	MIC	Interpretation	Antimicrobial	MIC	Interpretation
Cefoxitin Screen			Pradofloxacin	<= 0.12	S
Benzylpenicillin	>= 0.5	R	Inducible Clindamycin Resistance	NEG	-
Amoxicillin/Clavulanic Acid	<= 2	S	Erythromycin	>= 8	R
Oxacillin	<= 0.25	S	Clindamycin	0.25	S
Cefalotin	<= 2	S	Doxycycline	4	R
Cefpodoxime	<= 0.5	S	Minocycline	2	R
Cefovecin	<= 0.5	S	Nitrofurantoin	<= 16	S
Amikacin	<= 2	S	Chloramphenicol	>= 64	R
Gentamicin	<= 0.5	S	Florfenicol	<= 4	S
Enrofloxacin	<= 0.5	S	Trimethoprim/Sulfamethoxazole	>= 320	R
Marbofloxacin	<= 0.5	S			

AES Findings:	Last Modified: May 17, 2024 10:18 ICT Parameter Set: Global CLSI-based+Natural Resistance V2 17 NOV2017
Confidence Level:	Consistent

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1814-1 (Approved)

Card Type: AST-GP81 Bar Code: 3813162403170376 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity:

Selected Organism: Staphylococcus pseudintermedius

Action	Name (User ID)	Date/Time	Comment
--------	----------------	-----------	---------