

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1819-1 (Approved)

Card Type: AST-GN97 Bar Code: 6873220503566670 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity:

Selected Organism: Klebsiella pneumoniae

Comments:	

McFarland: 0.62 (0.50 - 0.63)

Identification Information	
Organism Origin	Technologist
Selected Organism	Klebsiella pneumoniae
	Entered: Dec 23, 2025 14:41 ICT By: LabTech
Analysis Messages:	

Card is for veterinary use only McFarland: 0.62 (0.50 - 0.63)

Susceptibility Information	Card: AST-GN97	Lot Number: 6873220503	Expires: Sep 20, 2026 12:00 ICT
	Status: Final	Analysis Time: 8.73 hours	Completed: Dec 23, 2025 23:27 ICT

Antimicrobial	MIC	Interpretation	Antimicrobial	MIC	Interpretation
ESBL	NEG	-	Gentamicin	>= 16	R
Ampicillin	>= 32	R	Neomycin	>= 64	R
Amoxicillin/Clavulanic Acid	>= 32	R	Enrofloxacin	>= 4	R
Cefalexin	>= 64	R	Marbofloxacin	>= 4	R
Cefalotin	>= 64	R	Pradofloxacin	>= 4	R
Cefpodoxime	>= 8	R	Doxycycline	>= 16	R
Cefovecin	>= 8	R	Tetracycline	>= 16	R
Ceftiofur	>= 8	R	Nitrofurantoin	128	R
Imipenem	2	I	Chloramphenicol	>= 64	R
Amikacin	>= 64	R	Trimethoprim/ Sulfamethoxazole	40	S

AES Findings:	Last Modified: May 17, 2024 10:18 ICT Parameter Set: Global CLSI-based+Natural Resistance V2 17 NOV2017
Confidence Level:	Consistent

bioMérieux Customer:

Laboratory Report

Printed by: LabTech

System #:

Isolate: 1819-1 (Approved)

Card Type: AST-GN97 Bar Code: 6873220503566670 Testing Instrument: 000019E1EF45 (VK2C 18325)

Setup Technologist: Laboratory Technician(LabTech)

Organism Quantity:

Selected Organism: Klebsiella pneumoniae

Action	Name (User ID)	Date/Time	Comment
--------	----------------	-----------	---------