

Shepard Test Stand - Milestone # 109: Test and Evaluate New Force Sensing Resistor

Status:	Closed	Priority:	Normal
Author:	Jeremy Wright	Category:	
Created:	01/16/2013	Assignee:	Jeremy Wright
Updated:	06/24/2013	Due date:	
Subject:	Test and Evaluate New Force Sensing Resistor		
Description			
A new flexible force sensor has been selected that has the proper force range. It needs to be integrated into the DAQ system and tested to see if it is adequate.			

History

01/19/2013 08:24 pm - Jeremy Wright

- Target version set to v1.1 Shepard Test Stand - Prototype

06/24/2013 06:53 am - Jeremy Wright

- Status changed from New to Closed

Aaron Harper found a load cell that was cheaper than the upgraded FSR we were going to use. The load cell worked well in testing and so the FSR design has been abandoned.