

KERASCAN Keratoconus Screening Report

Protocol Version: 2.0.0 | Generated: 2026-08-21 10:20 UTC

Screening Information

Field	Value
Screening ID	DEMO-2026-001
Age	16
Sex	Female
Site / School	St. Mary School
Screening Date	2026-08-21
Operator ID	OP-DEMO
Device ID	KERASCAN-001
Consent Recorded	Yes

Measurements

Measurement	OD (Right Eye)	OS (Left Eye)
K1 (flat K, D)	44.50	42.50
K1 Axis (°)	175	180
K2 / Steep K (D)	48.20	44.10
K2 Axis (°)	85	90
Kmax (D)	51.00	—
Mean K (D)	46.30	43.30
Pachymetry (µm)	462.00	535.00
Pachymetry Type	thinnest	central
Sphere (D)	-3.00	-1.25
Cylinder (D)	2.75	0.50
Cylinder Axis (°)	80	90
VA (logMAR)	—	—
Refraction Type	autorefraction	autorefraction
Quality	Good	Good

Per-Eye Screening Results

OD — Right Eye	
Engine result:	SUSPICIOUS
Decision:	PRIORITY_REFERRAL
Suspicious Rationale: Irregular astigmatism pattern, Elevated keratometry (K2), Low corneal thickness (pachymetry), High cylinder (astigmatism).	

OS — Left Eye	
Engine result:	NORMAL-LIKE
Decision:	SCREEN_NEGATIVE
Reason codes:	None

Overall Child-Level Decision

OVERALL RESULT	REFERRAL PRIORITY
PRIORITY_REFERRAL	PRIORITY

■ AI-assisted keratoconus screening result. This is not a confirmed diagnosis. Positive, discordant, ungradable or clinically concerning findings require repeat assessment, corneal tomography and qualified clinical review.

Image Analysis

OD — Original Image [Image not available]

OS — Original Image [Image not available]

Quality Findings

Eye	Gradable	Quality Score	Flags
OD	Yes	58.00	irregular_spacing
OS	Yes	74.00	None

ROI Detection

Eye	Method	Confidence	Radius (px)
OD	hough	0.81	188.00
OS	hough	0.94	193.00

Detailed Measurements & Repeat Readings

Eye	Reading #	K2 (D)	Pachy (μm)	Cyl (D)	Quality
OD	1	48.20	462.00	2.75	Good
OS	1	44.10	535.00	0.50	Good

Versions & Provenance

Item	Value
Protocol version	2.0.0
Pipeline version (OD)	phase1-0.1.0
Model hash (OD)	a1b2c3d4e5f6
Classification skipped (OD)	False
Prototype score (OD)	0.84

Decision Audit Trail

Level	Result	Overridden	Override By	Reason
child	PRIORITY_REFERRAL	No	—	—
eye	PRIORITY_REFERRAL	No	—	—
eye	SCREEN_NEGATIVE	No	—	—

Pentacam / Corneal Tomography Follow-Up

Field	Value
Exam Date	[Not yet recorded]
Kmax OD	
Kmax OS	
Belin-Ambrósio D OD	
Belin-Ambrósio D OS	
Performed By	
Notes	

AI-assisted keratoconus screening result. This is not a confirmed diagnosis. Positive, discordant, ungradable or clinically concerning findings require repeat assessment, corneal tomography and qualified clinical review.