KeaExpert release notes

Note: items in red may require modifications to your own sequences. All personal sequences should be recompiled after each software update.

Version 2.02.24 (09-November-2026)

- Release of Kea-NMR and NMR-Mouse (for PM systems) set of pulse programs

Version 2.02.20(08-October-2026)

- Allow wide textmenus or text boxes to have read-only strings
- Timing fixes in CPMG pulse programs for FX3
- Wider button was added as a new interface element for better distinguishing
- Further fixes and tidying up of pulse programs for PM systems
- Added/rename ServicePM as a macro to control the lift of PM systems

Version 2.02.18(18-August-2026)

- New 3.130 Prospa version with some new system commands and fixes
- Simulator can estimate pulse program duration
- Added libraries for FTDI
- Tidying up Macro folder
- Added the latest TRex firmware for FX3 Spartan 7
- Updated some procedures to be fully used for KeaExpert
- Added the possibility of importing double-precision data
- Updated documentation
- Minor fixes of batch mode procedures
- Added new menu entry "Open startup macro" to allow rapid access to local macros
- New 3.129 Prospa version with fixed cureditor commands
-

Version 2.02.17(09-July-2026)

- Implemented code for the trigger command
- Further porting of the Expert pulse programs for Kea systems
- Added some missing DLLs
- Added simulator procedure for several pulse programs
- Added new Kea icons and corresponding code

Version 2.02.16 (02-May-2026)

- Some minor fixes for simulator mode
- Fixed simulator mode required only one procedure within pp macro
- First version of PyQt interface ExpertController to run Expert
- Added necessary files to create Python environment and start ExpertController

Version 2.02.15 (21-April-2026)

- Added experiments 1PulseAmplitudeSweep and 1PulseDurationSweep.
- Added a simulation mode for certain experiments – see ‘Select Kea’ in the file menu to choose

this option.

Version 2.02.14 (21-March-2026)

- Ported experiments SSE and T1Sat from Mouse-Kea. Added T1Sat-ie which uses interexperiment time rather than repetition time.

Initial release Version 2.02.12 (13 February 2026)

- Port of 2.02.12 SpinsolveExpert to KeaExpert.
    - Removal of Spinsolve specific features.
    - Port of most important Kea-NMR and Kea-Mouse experiments.
    - Modification of preferences/common parameters interface to allow setting these values since for the Kea these may change more often than on the Spinsolve.
    - A new spectrometer parameter viewer/updater designed for the Kea.
    - Some experimental Mouse lift commands added