HEART RATE MONITOR

Manar Abdelatty

https://github.com/Manarabdelaty/HeartMonitor

Requirements

- Set ADC sample rate [150 SPS -- 1000 SPS]
- Collect ADC data for one minute
- Send ADC data over UART
- UART Baud rate at least (40 KHz)
- Display ECG signal real-time
- Calculate heart beats per minute

Block Diagram



Design


3.3V
USB-UART
DTR
RXD
TXD
+5V
GND
0xFF
0xEA
0x00
0xFF
16 bytes
B12
B13
B14
B15
A8
A9
A10
A11
A12
A15
B3
B4
B5
B6
B7
V3
G
STM 32
(Blue Pill)
B11
B10
B1
B0
A7
A6
A5
A4
A3
A2
A1
A0
B11
B11
B11
B11
B11
AD8282
GND
3.3V
OUTPUT
LO-
LO+
SDN`
GND
- Fixed Length 16-byte commands
  - Rate="{new_rate}";
  - Data;
  - hbpm;
- UART receive IT fired when all the 16-bytes are sent.


TIM3
USB-UART
DTR
RXD
TXD
+5V
GND
rate={new_rate};
AD8282
GND
3.3V
OUTPUT
LO-
LO+
SDN`
GND
B12
B13
B14
B15
A8
A9
A10
A11
A12
A15
B3
B4
B5
B6
B7
V3
G
TIM3
STM 32
(Black Pill)
B11
B10
B1
B0
A7
A6
A5
A4
A3
A2
A1
A0
B11
B11
B11
B11
B11
3.3V
1) Sample Rate Command
rate={new_rate};
ARR = Sysclk / pre-scaler / new_rate
Prescaler
3999
Sysclk
8MHz
ARR
??
TIM3


3.3V
USB-UART
DTR
RXD
TXD
+5V
GND
data;
AD8282
GND
3.3V
OUTPUT
LO-
LO+
SDN`
GND
B12
B13
B14
B15
A8
A9
A10
A11
A12
A15
B3
B4
B5
B6
B7
V3
G
TIM3
TIM2
STM 32
(Black Pill)
ADC
B11
B10
B1
B0
A7
A6
A5
A4
A3
A2
A1
A0
B11
B11
B11
B11
B11
2) Collect Data Command “data;”
- start TIM3 to trigger ADC conversion
- start TIM2 to count for one minute
- stop two timers after one minute elapses
TIM2 Counter Period



3) Compute heartbeats per minute
Command “hbpm;”

- **start TIM3 to trigger ADC conversion**
- **start TIM2 to measure distance between two peaks.**
- **stop two timers after calculating the heart rate**

Heart Rate Calculation

```
Bpm = 60 / peak_distance
```





Peak Detection → Peak Threshold → Low Pass Filter in SW → Distance between two R-Peaks

3

SW Archietecutre

Round Robin with Interrupts

- **UART_Receive_Interrupt**
  - Receives commands from the Python application.
- **TIM2_Period_Elapsed_Interrupt**
  - Generated when one minute passes
- **ADC_Conversion_Interrupt**
  - Generated on one complete ADC conversion

```c
void HAL_UART_RxCpltCallback(UART_HandleTypeDef* huart){
  if(huart->Instance == USART1){
      decode();  // first decode instruction
      HAL_UART_Receive_IT(&huart1,
      (uint8_t *)rxBuffer,
      16);
  }
}
```

```c
void decode(){
  // fixed length commands
  char command [COMMAND_LENGTH+1] = {rxBuffer[0],
  rxBuffer[1], rxBuffer[2], rxBuffer[3], rxBuffer[4], '\0'};
  if(strcmp(command, "rate=") == 0){
    // set new sample rate
    char sample_rate_str [MAX_UART];
    for (int i= COMMAND_LENGTH; i<MAX_UART; i++){
    set_sample_rate = 1;
    new_sample_rate = atoi(sample_rate_str);
  } else if (strcmp(command, "hbpm;") == 0){
    // compute heart beats
    compute_bpm = 1;
  } else if (strcmp(command, "data;") == 0){
    // collect one minute of data
    collect_data = 1;
  }
  else {
    // Invalid command
  }
}
```

UART_Receive_Interrupt

```
27 void HAL_TIM_PeriodElapsedCallback(TIM_HandleTypeDef *  htim){
28   if(htim->Instance == TIM2){
29     reset = reset + 1;
30     if(__HAL_TIM_GET_FLAG(&htim2, TIM_FLAG_CC1) != RESET){
31       timer_done = timer_done + 1;
32       HAL_TIM_Base_Stop(&htim3);       // stop ADC trigger timer
33       HAL_TIM_Base_Stop_IT(&htim2);  // stop one minute timer
34     }
35   }
36 }
```

TIM2_Period_Elapsed_Interrupt

```
void HAL_ADC_ConvCpltCallback(ADC_HandleTypeDef* hadc){
  if(hadc->Instance == ADC1){
    adc_value = HAL_ADC_GetValue(&hadc1);
    if (transmit_adc){ // if data command triggered ADC
      sprintf(value, "%d\n\r", adc_value);
      HAL_UART_Transmit(&huart1, (uint8_t *)value, strlen(value), 10);
    }else if(detect_peak){ // if hbpm command triggered ADC
        adc_values[count] = adc_value;
        if(count <= 2) {
          countPeaks();
        }
        count = (count + 1) % MIN_SAMPLES_RR;
    }
    sample_count++;
  }
}
```

ADC_Conversion_Complete_Interrupt

```
void countPeaks(){
  if((adc_values[count-1] > adc_values[count-2]) &&
     (adc_values[count-1] > adc_values[count])){    // if the previous point is a l
       if(adc_values[count-1] > PEAK_THRESHOLD ){  // if the previous point is abov
            curr_peak = __HAL_TIM_GET_COUNTER(&htim2) / 1000.0; // get current tir
            peak_distance = curr_peak - prev_peak;              // Compute RR-Inte
            if(peak_distance > MIN_RR_INTERVAL){                // if the RR-Inter
                num_peaks++;                                    // count it as a p
                if (num_peaks >= 2) {
                  computed_bpm = 60.0 / (peak_distance);         // 300-method for
                  transmit_bpm = 1;
                  sprintf(bpm, "bpm: %d\n\r", computed_bpm);
                  HAL_UART_Transmit(&huart1, (uint8_t *)bpm, strlen(bpm), 10);
                }
                prev_peak = curr_peak;
            }
       }
    }
}
```

Count Peaks

```c
if(collect_data){
  sample_count = 0;
  transmit_adc = 1;
  // start TIM3 & TIM2 for one minute
  HAL_TIM_Base_Start(&htim3);
  HAL_TIM_Base_Start_IT(&htim2);

  __HAL_UART_DISABLE_IT(&huart1, UART_IT_RXNE);
  collect_data = 0;
  __HAL_UART_ENABLE_IT(&huart1, UART_IT_RXNE);
}
```

```c
if (compute_bpm) {
  // start adc sample rate timer; it is stopped
  detect_peak = 1;

  HAL_TIM_Base_Start(&htim3);
  HAL_TIM_Base_Start_IT(&htim2);

  __HAL_UART_DISABLE_IT(&huart1, UART_IT_RXNE);
  compute_bpm = 0;
  __HAL_UART_ENABLE_IT(&huart1, UART_IT_RXNE);

}
```

```c
if(set_sample_rate){
  // change ARR value of TIM3
  new_counter_period = ((float) counter_clk / (float)new_sample_rate) - 1;
  __HAL_TIM_SET_AUTORELOAD(&htim3, new_counter_period);

  __HAL_UART_DISABLE_IT(&huart1, UART_IT_RXNE);
  set_sample_rate = 0;
  __HAL_UART_ENABLE_IT(&huart1, UART_IT_RXNE);

}
```

Main Loop

Python Application

- **Pyserial**
  - UART communication
- **Matplotlib Animation**
  - Real-time plotting
- **tkinter**
  - GUI



Available ports list

Opens the selected port.

Sends the rate command

Opens a new page to collect one minute of data

Main Page

Real-Time Display

DEMO