



CALIBRATION CERTIFICATE

Customer:	MW RICE & SHINE	Particulars:	Sphygmomanometer
Address:	AGUSAN DEL NORTE, BUTUAN CITY, AMBAGO	Model:	Standard Model
Reference No.:	REF-20250929-0001	Serial No.:	MET - 2563
Sample No.:	MET - 2563	Range:	Ordinary Accuracy III (NAWI)
Date Submitted:	29 Sep 2025	Graduation:	1 mmHg
Date Calibrated:	29 Sep 2025		
Place of Calibration:	DOST Regional Office No. I - RSTL		

TEST RESULTS

I. Test for the Maximum Permissible Errors of the Cuff Pressure Indication

Applied Pressure (mmHg)	Sample Reading (mmHg)		Maximum Deviation	Maximum Permissible Error
	Increasing	Decreasing		
0	0	0	0	within $\hat{A} \pm 4$ mmHg
50	51	50.8	50	within $\hat{A} \pm 4$ mmHg
100	99.8	99.6	100	within $\hat{A} \pm 4$ mmHg
150	148.5	148.4	150	within $\hat{A} \pm 4$ mmHg
200	198.3	198.3	200	within $\hat{A} \pm 4$ mmHg
250	248.8	248.8	250	within $\hat{A} \pm 4$ mmHg
300	298.5	298.7	300	within $\hat{A} \pm 4$ mmHg

II. Test for Hysteresis Error

Maximum Hysteresis Error	Maximum Permissible Error
0 mmHg	within $\hat{A} \pm 4$ mmHg

III. Test for Air Leakage of the Pneumatic System

Applied Pressure (mmHg)	Sample Reading (mmHg)		Rate of Pressure Loss	Maximum Permissible Error
	1st Reading	After 5 minutes		
300	60	60	0 mmHg/min	$\hat{a} \% \pm 4.0$ mmHg/min
250	120	120	0 mmHg/min	$\hat{a} \% \pm 4.0$ mmHg/min
200	180	180	0 mmHg/min	$\hat{a} \% \pm 4.0$ mmHg/min
150	250	238	2.4 mmHg/min	$\hat{a} \% \pm 4.0$ mmHg/min
100	300	296	0.8 mmHg/min	$\hat{a} \% \pm 4.0$ mmHg/min

IV. Test for the Rapid Exhaust Valve

Applied Pressure (mmHg)	Time for the Pressure Reduction to reach $\hat{a} \% \pm 15$ mmHg	Maximum Permissible Error
300		< 10 seconds

This certifies that the above Non-invasive Sphygmomanometer was Tested by the Regional Metrology Laboratory at Government Center, Sevilla, City of San Fernando, La Union.



The sphygmomanometer was tested according to its normal use with the pressure inlet connection at the bottom.
Testing range: from 0 mmHg to 300 mmHg positive gauge pressures.

Environmental conditions during testing:

Ambient Temperature ($^{\circ}\text{C}$): 23.0 $^{\circ}\text{C}$
Relative Humidity (RH): 45 %

The sphygmomanometer was subjected to air leakage, rapid exhaust valve, cuff pressure indication, and hysteresis tests.

Testing was done in accordance with the laboratory's in-house calibration method CM-MET-008 "Performance Testing of Sphygmomanometer" based on the International Recommendation OIML R 16-1 Edition 2002 "Non-invasive Mechanical Sphygmomanometers".

STANDARD USED AND TRACEABILITY:

Standard Gauge

Type:	Digital Pressure Calibrator
Model/Maker:	ADT672-05-GP15-BAR-N
Measurement Range:	0-1 bar (0-750 mmHg)
Accuracy (% FS):	0.05
Serial No.:	2731706006
Certificate No.:	102023-INS-08090
Traceability:	Metal Industry Research and Development Center
Calibration Date:	Oct-25

REMARKS:

- The general condition and workmanship of the instrument are SATISFACTORY
- The results given in this report were obtained at the time of the test and refer only to the particular sphygmomanometer submitted under the conditions indicated. The certificate is not intended to be representative of similar items. If the unit is modified or damaged in anyway, the results may be rendered invalid and will require re-verification.
- The user should determine suitability of the sphygmomanometer for its intended use.
- No adjustments were performed on the sphygmomanometer.
- The sphygmomanometer reading may be corrected by applying a correction of 1 mmHg. (e.g. Sphygmomanometer Reading: 120/80; After correction: 121/81, BASED ON THE CUFF PRESSURE INDICATION)
- The user is obliged to have the unit recalibrated at appropriate intervals.
- This report shall not be reproduced except in full, without the written approval of the laboratory.

Calibrated by:

MA. FERNANDA I. BANDA

Calibration engineer

Certified by:

Alot

Technical Manager