

Operational Qualification HeartPulse Android and iOS - Messy No IDs

This OQ document is intentionally messy. There are no OQ identifiers. The application should auto-generate AUTO-OQ IDs and map by meaning.

Verify heart rate capture using camera or wearable sensor. Expected result the app records bpm value, source and timestamp. Verify manual heart rate entry when sensor reading is not possible. Expected result the user can save manual bpm with reason and timestamp. Verify abnormal alert when bpm is above high threshold or below low threshold. Expected result warning message and alert log are created. Verify history screen displays saved readings with date time bpm source and trend chart. Expected result history is accurate for the logged in user. Verify cloud synchronization between Android and iOS account. Expected result reading created on one platform appears on the other platform after sync. Verify unauthorized user cannot view another user heart rate data. Expected result access is blocked and only own records are visible. Verify privacy notice and consent prompt appear before first data storage. Expected result user consent is recorded before saving data. Verify report export with date range as PDF and CSV. Expected result report file downloads and contains selected heart rate readings.