CINDER PEAK OBSERVATORY

INSTRUMENT DESK — SENSOR CALIBRATION SHEET

ISSUED 22 APRIL 2026 — CHECKED BY INES VARGA

SENSOR	OFFSET
DOME HUMIDITY PROBE	0.12
MIRROR CELL THERMISTOR	0.35
PRIMARY FOCUS ENCODER	0.08

SENSOR	OFFSET
SEEING MONITOR CAMERA	0.21
WIND MAST ANEMOMETER	0.47
SKY QUALITY METER	0.16

Calibration notes

Each offset on the previous page is subtracted from the sensor’s raw reading before the value enters the nightly log. The six sensors were checked against the instrument desk’s reference set on 21 April 2026, and the sheet replaces every earlier calibration sheet for these sensors. The next scheduled calibration is due in October 2026, or sooner if any sensor is moved or replaced. Questions about the sheet go to the instrument desk.