

Tool Validation Record

Tool Name

Selenium WebDriver

Tool Version

2.42

CM Tag

Manufacturer

Open Source

Intended Use

See over

continue over
if required

GAMP / cGMP Validation Category

☐ Category 1 / 2

☒ Category 3

☐ Category 4

☐ Category 5

Tool Validation Needs Assessment

Q1. Will the output of the tool be used to eliminate, reduce, or automate part of medical device production or an objective of the Quality System?

☒ YES

☐ NO

Q2. Will the tool's output **not** be verified or confirmed by other verification activities as specified in the Quality System?

☒ YES

☐ NO

Q3. Can the tool insert an error into the medical device sw / hw or fail to detect an existing error in the sw / hw within the scope of its intended usage?

☒ YES

☐ NO

Risk Assessment

DHFR No.

Title

Initial Analysis

8038

Selenium WebDriver 2.42 Risk Analysis

Initial RPN
(most severe)

☐ UNACCEPTABLE

☐ marginally tolerable

☐ TOLERABLE

☒ ACCEPTABLE

Q. Does the tool have electronic records?

YES

☐

NO

☒

Q. Does the tool use electronic signatures?

☐

☒

Validation Planning

Validation Activities

Validation Report

Title: Tool Validation

Subtitle: Validation Needs Assessment Form

DHFR: 119

Version: A01

Author: Technical Specialist

Approver: Functional Head



Date: ____ / ____ / ____

Date: ____ / ____ / ____

Intended Use
(continued if reqd.)

Selenium WebDriver is an Test Automation Library and the intended use is to:

- locate the web elements on the GUI of the web application by the automated test script
- interact with the web elements on the screen i.e. clicking, entering values etc. by the automated test script
- Get the web element information from GUI to the automated test script for test analysis

GAMP / cGMP Categories

Category 1 / 2
Infrastructure Software

This category is broken down into two subcategories: (1) established or commercially available layered software and (2) infrastructure software tools. Examples of the first subcategory are: databases, middleware, office software. Examples of the second subcategory are: antivirus, backup, network monitoring software.

Category 3
Nonconfigured Products

This category is for software that is capable of operating and automating business processes out-of-the-box, or that only requires run-time configuration to enable the software to operate within the installed environment. Examples of this category are: commercial-off-the-shelf (COTS), operating systems, or lab instruments / software

Category 4
Configured Products

This category is for software that must be modified to match a business process i.e. the user has the means and knowledge to change the functionality of the device in a way that changes the results output by the device. Examples of this category are: LIMS, SCADA, DCS, CDS, etc.

Category 5
Custom Applications

This category is for software that is customised (vs. configured) i.e. the application is developed using custom code for (or within) the application. Examples of software in this category are: bespoke software used in Leica Production such as DMS, PSS.



GAMP 5 Appendix M4(1) notes “custom software components..., developed with an internal scripting language, written or modified to satisfy specific user business requirements should be treated as Category 5”.