

	Leica Biosystems Melbourne Department: QA/RC
<i>Title:</i> Regulatory Change Assessment Form	
Change Request: 12773	
Product Name: BOND-III and BOND-MAX	

DOCUMENT APPROVAL

<i>Author:</i>	Signed by: <i>Anisha Behl</i>  Signer Name: Anisha Behl Signing Reason: I am the author of this document Signing Time: 18-Jun-2025 21:14:46 PDT 21454E0326BA44B79861DF59A8D7C0F7 <i>ASSOCIATE QA/RC</i>
<i>Reviewer and Approver:</i>	Signed by: <i>Gabriela Morales Mendez</i>  Signer Name: Gabriela Morales Mendez Signing Reason: I approve this document Signing Time: 19-Jun-2025 15:28:44 AEST C622DF48833E4221BEE0317E9B0529F9 <i>Senior Specialist, RC</i>

REGULATORY CHANGE ASSESSMENT FORM

SECTION 1: Change Request Details		
1.1	Change Request #	CR12773
1.2	Date of Assessment:	19 June 2025
1.3	Product Name:	<u>USA 510(k):</u> • Progesterone Receptor PGR Clone 16 • Estrogen Receptor Clone 6F11 (ER 6F11) • BOND MMR Antibody Panel <u>USA PMA:</u> • BOND ORACLE HER2 IHC SYSTEM <u>USA Class I 510(k) exempt:</u> • BOND-III • BOND-MAX <u>Canada and rest of world:</u> • BOND-III • BOND-MAX <u>Canada Class III:</u> Leica HER2 FISH System (on BOND-III and BOND-MAX)
1.4	Device Description:	<u>510 (k):</u> • PGR Clone 16 is a monoclonal mouse antibody that detects a human progesterone receptor epitope located in the nucleus of PGR positive cells. • ER 6F11 is a monoclonal mouse antibody that detects a human estrogen receptor epitope located in the nucleus of ER positive cells. • BOND MMR Antibody Panel detects the mismatch repair protein deficiency as an aid in the identification of HPNCC/Lynch Syndrome in patients diagnosed with CRC. <u>PMA:</u> • The Bond Oracle HER2 IHC System uses a mouse monoclonal antibody (Clone CB1 1) for the in vitro semi-quantitative detection of c-erb-2 antigen in sections of formalin-fixed, paraffin-embedded normal and neoplastic tissue. Visualization of the targeted antigen is achieved by sequential application of the HER2 Primary Antibody, visualization reagents, and chromogen, resulting in a visible reaction at the antigen site. Results are evaluated using a light microscope. <u>Canada Class III:</u> • For use on Leica Biosystems' Bond-MAX System and BOND-III System. TA9217 is a fluorescence in situ hybridization product designed to stain 30 tests (30 slides stained with LSI HER2/CEP17 Dual Probe). <u>Canada Class II/US Class I/Rest of world:</u> The BOND system is designed for immunostaining histological specimens mounted on microscope slides. The BOND System intended use is: The BOND system automates clinical protocols for immunostaining of pathology specimens mounted on microscope slides. Microscope slides subsequently undergo interpretation by a qualified healthcare professional to aid diagnosis.

REGULATORY CHANGE ASSESSMENT FORM

1.5	Description of Change(s):	Refurbished BOND-III and BOND-MAX instruments to be sold in the countries mentioned in Appendix A are within scope of this change. BOND-III and BOND-MAX can be repaired and refurbished. This change is to generate new part numbers (RS21.2201 and RS49.0051) for BOND-III and BOND-MAX refurbished instruments. As part of this change the following labels will be released: <table border="1" data-bbox="507 562 1189 813"> <tr> <td>21.6602</td><td>BOND-III Pre-Owned Rating Plate Label</td></tr> <tr> <td>21.6603</td><td>BOND-III Pre-Owned Outer Pack Label</td></tr> <tr> <td>49.6105</td><td>BOND-MAX Pre-Owned Rating Plate Label</td></tr> <tr> <td>49.6106</td><td>BOND-MAX Pre-Owned Outer Pack Label</td></tr> </table> The new rating plates will include Chinese translations. The content/symbols of the label remain the same compared to the one of a new instrument, with the only difference being: • Reference to the new part number RS 21.2201 • New UDI number for BOND-III: 09349458005672 • Reference to the new part number RS 49.0051 • New UDI number for BOND-MAX: 09349458005689 The proposed change is to the instrument numbering (Product Code), with no changes to the instrument functionality. The changes made do not impact performance and/or the staining function of the instrument and its intended use; and does not introduce any risk to the instrument specifications or build process. There is also no impact to the safety of the instruments. As part of this change, we are updating the technical file to include the new product codes. NOTE: A separate regulatory assessment will be necessary for markets not included in Appendix A prior to the sale of refurbished instruments in those markets.		21.6602	BOND-III Pre-Owned Rating Plate Label	21.6603	BOND-III Pre-Owned Outer Pack Label	49.6105	BOND-MAX Pre-Owned Rating Plate Label	49.6106	BOND-MAX Pre-Owned Outer Pack Label
21.6602	BOND-III Pre-Owned Rating Plate Label										
21.6603	BOND-III Pre-Owned Outer Pack Label										
49.6105	BOND-MAX Pre-Owned Rating Plate Label										
49.6106	BOND-MAX Pre-Owned Outer Pack Label										
1.6	Reason for Change(s):	To generate a new instrument number for the refurbished instruments and to update the labels accordingly.									
1.7	Product registration status	N/A ☒	Leica Biosystems Melbourne currently does not manufacture any US Class I IVDs where a 510(k) is applicable. U.S Class II 510 (k) exists, complete Section 2 U.S 510 (k) # K193393, BOND Ready-to-Use (RTU) Primary Antibody Progesterone Receptor (16); Novocastra Liquid Mouse Monoclonal Antibody Progesterone Receptor Clone 16, 510 (k) Holder: Leica Biosystems Newcastle Ltd, Date of Clearance: 06 March 2020 U.S 510 (k) #: K183100, ESTROGEN RECEPTOR CLONE 6F11 (ER 6F11)								

REGULATORY CHANGE ASSESSMENT FORM

		510 (k) Holder: Leica Biosystems Newcastle Ltd, Date of Clearance: 07 December 2018 U.S 510 (k) #: K213348, BOND MMR Antibody Panel 510 (k) Holder: Leica Biosystems Newcastle Ltd, Date of Clearance: 21 February 2023
	☒	U.S Class III PMA exists, complete Section 3 PMA # P090015, BOND ORACLE HER2 IHC SYSTEM PMA Holder: Leica Biosystems Newcastle Ltd Date of approval: 18 Apr 2012 Last supplement # S012
	☒	Product is U.S 510(k) Exempt, complete Section 4
	☒	Product is Class II in Canada, complete Section 5 Device license # 66466 (BOND-MAX System) Device license # 81061 (BOND-III System)
	☒	Product is Class III or Class IV in Canada, complete Section 6 Device license # 89629, Class III (BOND-MAX System) Device license # 89628, Class III (BOND-III System)
	☐	Product is Class I in Brazil, complete Section 7
	☒	Product is Class I in Japan, complete Section 8 Device license # 13B2X10268BM0001 (BOND-MAX) Device license # 13B2X10268B30001 (BOND-III)
	☒	Product is Class I in Australia, complete Section 9 Device license # 176143
	☒	Product is registered in regions outside of US, complete Section 10

REGULATORY CHANGE ASSESSMENT FORM

SECTION 2: Registration Impact Assessment [Premarket Notification, 510(k)]

2. i.	Is this a New Product?	☐ Yes	☒ No, continue to 2. ii.
	<i>If Yes, provide details of the U.S FDA Product Registration Classification</i>		
2. ii.	Is this a change due to recall or corrective action?	☐ Yes	☒ No, continue to 2. iii.
	<i>If Yes, provide details (Follow U.S FDA ODE Bluebook Memorandum K95-1 (Nov 21, 1995), "510(k) requirements during Firm-Initiated Recalls")</i>		
2. iii.	Is this a change made with intent to significantly improve the safety or effectiveness of the device?	☐ Yes	☒ No, continue to 2. iv.
	<i>If Yes, New 510 (k) is required</i>		
2. iv	Is this a Labelling change?	☒ Yes, continue to 2.A	☐ No, continue to 2.v <i>Section 2.A may be deleted</i>
2. v	Is this a Technology, Engineering, or Performance Change (for a non-IVD)?	N/A, continue to 2. vi. <i>Leica Biosystems Melbourne instruments are classified as IVD's.</i>	
2. vi	Is this a Materials Change (for a non-IVD)?	N/A, continue to 2. vii. <i>Leica Biosystems Melbourne instruments are classified as IVD's.</i>	
2. vii	Is this a Technology, Engineering, or Performance Change <u>for an IVD</u> ?	☒ Yes <i>continue to 2.B</i>	☐ No, continue to 2. viii <i>Section 2.B may be deleted</i>
2. viii	Is this a Software or Firmware Change?	☐ Yes <i>continue to 2.C</i>	☒ No, continue to next applicable section. <i>Section 2.C may be deleted</i>

2.A	Labelling Change		
2. A1.	Is it a change in the indications for use statement?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, the change is not for indication for use statement		
2. A1.1	Is it a change from a device labelled for single use only to a device labelled as reusable?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, the change is not from a device labelled for single use only to a device labelled as reusable		
2. A1.2	Is it a change from prescription (Rx) to over the counter (OTC) use?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, the change does not change from prescription (Rx) to over the counter (OTC) use		
2. A1.3	Is it a change to the device name or a change solely to improve readability or clarity?	☐ Yes <i>If Yes, Documentation</i>	☒ No
	<i>Detailed justification:</i> No, the change is not to device name or a change solely to improve readability or clarity		
2. A1.4	Does the change describe a new disease, condition or patient population that the device is intended for use in diagnosing, treating, preventing, curing or mitigating?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, the change does not describe a new disease, condition or patient population that the device is intended for use in diagnosing, treating, preventing, curing or mitigating		

REGULATORY CHANGE ASSESSMENT FORM

2.A	Labelling Change		
2. A1.5	Does a risk-based assessment of the changed device identify any new risks or significantly modified existing risks?	☐ Yes <i>If Yes, a new 510(k) may be required. (if only adding contraindication, submit CBE 510(k))</i>	☒ No <i>If No, Documentation</i>
	<i>Detailed justification:</i> No new risk is introduced as a result of this change.		
2. A2.	Does the change add or delete a contraindication?	☐ Yes <i>If Yes, a new 510(k) may be required. (if only adding contraindication, submit CBE 510(k))</i>	☒ No <i>If No, Documentation</i>
	<i>Detailed justification:</i> No contraindication was added or deleted		
2. A3.	Is it a change in warnings or precautions?	☐ Yes	☒ No <i>If No, Documentation</i>
	<i>Detailed justification:</i> There is no change to precaution or warnings		
2. A4.	Could the change affect the directions for use of the device?	☐ Yes	☒ No <i>If No, Documentation</i>
	<i>Detailed justification:</i> The change does not affect the directions for use of the device		

2.B	Is this a Technology, Engineering, Performance, and Materials Change for In Vitro Diagnostic Devices (IVDs)?		
2. B1.	Does the change alter the operating principle of the IVD?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> The change does not alter the operating principle of the IVD		
2. B2.	Is the change identified in a device-specific final guidance or classification regulation?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, the change is not identified in a device-specific final guidance or classification regulation.		
2. B3.	Does a risk assessment of the changed device identify any new risks or significantly modified existing risks?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No new risk or significant modifications to existing risks identified. Refer to supporting documents in section 11.3.		
2. B4.	Do design verification and/or validation activities produce any unexpected issues of safety or effectiveness?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> Design verification and/or validation activities did not produce any unexpected issues of safety or effectiveness		

REGULATORY CHANGE ASSESSMENT FORM

SECTION 3: Registration Impact Assessment [Premarket Approval, PMA]		
3.A	Identify the modification and reason for the change	
	Refer to section 1.5	
3.B	Reference and/or attach the Risk analysis performed for the proposed change	
	Refer to section 11.3.	
3.C	Document the assessment of impact due to modification on safety & effectiveness (S&E)	
	No impact to safety or effectiveness.	
3.D	180-Day PMA Supplement	☐ Yes <i>If Yes, a 180-Day PMA Supplement is required.</i>
	☒ No Go to 3.E	
	Is this a significant change requiring an in-depth review and FDA approval prior to implementation of change? The changes detailed below are examples of significant changes • Changes to the principle of operation; • Changes to control mechanism; • Changes to device design or performance; • Changes to labeling; • New testing requirements or acceptance criteria. <i>Detailed justification:</i> The change is not related to any of the above.	
3.E	Real-Time Supplement	☐ Yes <i>If Yes, a Real-Time Supplement is required.</i>
	☒ No Go to 3.F	
	Is this is a minor change in device design, software, sterilization and packaging, labeling (i.e. instructions for use, warning or precaution or other labeling that does not affect the indications or contraindications)? <i>Detailed justification:</i> The change is in the labelling, to add a new part numbers for refurbished instruments.	

REGULATORY CHANGE ASSESSMENT FORM

3.F	Special PMA Supplement	☐ Yes <i>If Yes, a Special PMA Supplement is required.</i>	☒ No Go to 3.G
Is this a labeling and manufacturing changes that enhance the safety of the device or the safety in the use of the device AND has no impact on device effectiveness (e.g. new and or modified label to add or strengthen a contraindication, warning, precaution, or information about an adverse reaction; an added step to the quality control or manufacturing processes to enhance safety only). The changes detailed below are examples of Special PMA Supplement • New warning added; • Manufacturing Process change; • Improved instructions; and • Additional Inspection step in the manufacturing process; <i>Detailed justification:</i> Not applicable. No new warnings or precautions added.			
3.G	30-Day Notice	☐ Yes <i>If Yes, a 30-Day Notice is required.</i>	☒ No Go to 3.H
Is this a change to manufacturing procedures or method of manufacture to any materials that are critical to the performance of the device, thereby affecting safety or effectiveness of the device? The changes detailed below are examples of 30-Day notice: • Sterilization Process Parameters (e.g. sterility dose auditing, time of sterilization) • Automating existing processes (e.g. manual to automated cutting process, manual-controlled machine to computer-controlled machine) • Joining Processes (e.g. bonding equipment to replace existing equipment with a different model heat bonder. Note: A validated change that replaces the equipment with identical equipment using the same parameters as originally validated will not require a 30-day.) • Cleaning Methods Used to Remove Manufacturing Materials (e.g. changing wash process to increase efficiency of process and reduce cost) • Changes to Manufacturing Materials • Environmental Conditions for Manufacturing, Storage or Distribution Facilities (e.g. change to add new cleanroom to existing manufacturing space, relocation of formulation room within facility, change in ISO class. Note: This is not similar to change in establishment or facility where manufacturing, processing or packaging of the device occurs.) • Suppliers of Critical Components, Materials or Services, where specifications are unchanged (e.g. using an additional supplier for welding of components, adding a raw material supplier, adding an alternate qualified supplier, making a change in sterilization test site for sterilization of test samples. Note: A 30-day is not appropriate where any manufacturing change which involves adding a second supplier of a critical component has different specifications. This would be a 180-day or real-time review) • Quality Control Testing (e.g. changes to quality control testing used on incoming components, raw materials, the in-process or finished device, performing end-product testing prior to sterilization, eliminating certain in-process or final device test to eliminate redundancy, waste or creating efficiency, changing sample size in an assay, removing rejection criterion for component acceptance within an in-process inspection step, adding inspection steps in response to field failures of the device) Inadequate information and supportive data can lead to conversion of 30-Day to 135-Day Supplement. It must be identified if testing was driven by the risk assessment and if it was additional testing, and provide an explanation for why any additional testing conducted was voluntary and not required by the quality system. <i>Detailed justification:</i> No impact to the safety and effectiveness of the instrument.			

REGULATORY CHANGE ASSESSMENT FORM

3.H	Annual (Periodic) Report	☒ Yes <i>If Yes, consider including in Annual (Periodic) Report.</i>	☐ No only if there is no PMA
	Is the change implemented to manufacturing procedure or method of manufacturing that does not affect safety or effectiveness of the device, are functionally equivalent, have not changed any specifications and do NOT require any other submission type (i.e. validation data required to support change)? Examples include but are not limited to: • Editorial changes in labeling or SOP to make description clearer; • Reports in scientific literature concerning the device; • Unpublished reports of clinical and non-clinical investigations associated with device; • Redundant controls/checks; • Addition of alternative supplier for component with same specification; • Extension of expiration date of the device based on data obtained under an approved protocol; • Change in temperature to provide comfortable environment for personnel; • Increasing number of products sampled for release testing; • Tightening viscosity range; • Adding step for clarification to provide additional assurance of device quality; and • Updating a process to include a backup.		
	<i>Detailed justification:</i> No impact to safety and effectiveness		

REGULATORY CHANGE ASSESSMENT FORM

SECTION 4: Assessment of Change – 510(k) exempt devices

Is this a change to a 510(k) exempt device?		☒ Yes	☐ No <i>Go to 5</i>
4.1	Does the change modify the device intended use and/or indications for use?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, the change does not modify the device intended use and/or indications for use		
4.2	Is it a change to the intended device users (e.g. a change to lay users from health care professional users)?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, there is no change to the intended device users		
4.3	Is it a change to the device operating principle that changes the fundamental scientific technology of the device (e.g. a change to the staining method)?	☐ Yes <i>If Yes, a new 510(k) may be required.</i>	☒ No
	<i>Detailed justification:</i> No, there is no change to the device operating principle that changes the fundamental scientific technology of the device		
4.4	Does the change result in a new device classification code or name?	☐ Yes	☒ No
	<i>Detailed justification:</i> No, the change does not result in a new device classification code or name		
4.5	Is it a change to the device proprietary name, addition of a new proprietary name or other change to device information listed in the FURLS (FDA Unified Registration and Listing system) DRLM (device Registration and Listing Module)?	☐ Yes	☒ No
	<i>Detailed justification:</i> No, there is no change to the device proprietary name, addition of a new proprietary name or other change to device information listed in the FURLS		
4.6	Does this change impact the UDI-DI?	☒ Yes	☐ No
	<i>Consider if an update is required to the information registered with GUDID:</i> • New Issuing Agency • Changes to quantity of devices • Product is supplied as a kit • Product name change • New version or model		
	<i>Detailed justification:</i> UDI-DI information was updated, since this change is to generate new part numbers for the refurbished instruments.		

REGULATORY CHANGE ASSESSMENT FORM

SECTION 5: Assessment of Change – Canada II

5.A	Is this a change to a Canada Class II device?	☒ Yes	☐ No <i>Go to 6</i>
5.A1	Does the change affect the classification of the device? <i>Detailed justification:</i> This change does not affect the device classification.	☐ Yes	☒ No
5.A2	Is it a change to the name of the manufacturer? <i>Detailed justification:</i> No change to the name of the manufacturer.	☐ Yes	☒ No
5.A3	Is it a change to the name of the device? <i>Detailed justification:</i> No change to the name of the device.	☐ Yes	☒ No
5.A4	Is it a change to the identifier of the device, including the identifier of any medical device that is part of a system, test kit, medical device group, medical device family? <i>Detailed justification:</i> Two new codes are being created as part of this change. Refer to section 1.5.	☒ Yes	☐ No
5.A5	Is it a change in the medical conditions, purposes or uses for which the device is manufactured, sold or represented? <i>Detailed justification:</i> This change does not impact the intended purpose of the device.	☐ Yes	☒ No
5.A6	Does this change impact compliance with Canada Safety and Effectiveness and Labelling requirements? <i>Detailed justification:</i> No impact to the compliance with Canada Safety and Effectiveness and Labelling requirements.	☐ Yes	☒ No

5.B	Summary		
5.B1	Are answers to any of the above questions YES? <i>Note: An application for an amended medical device license for Class II devices should be submitted.</i>	☒ Yes <i>Amendment required</i>	☐ No <i>Go to 6</i>

REGULATORY CHANGE ASSESSMENT FORM

SECTION 6: Assessment of Change – Canada Class III or IV

6.A	Is this a change to a Canada Class III or IV device?	☒ Yes	☐ No <i>Go to 7</i>
6.A1	Is the change due to a recall?	☐ Yes	☒ No
	<i>Detailed justification:</i> This change is not driven by a recall.		
6.A2	Is this a change to the manufacturing processes, facility or equipment?	☐ Yes <i>Go to 6.B</i>	☒ No
	<i>Detailed justification:</i> No change to the manufacturing process, the refurbished instruments undergo commissioning, chemical testing, decontamination, and packing as all other new instruments.		
6.A3	Is this a change to the manufacturing quality control procedures?	☐ Yes <i>Go to 6.C</i>	☒ No
	<i>Detailed justification:</i> No change to the manufacturing quality control procedures.		
6.A4	Is this a change in design?	☐ Yes <i>Go to 6.D</i>	☒ No
	<i>Detailed justification:</i> No change in design.		
6.A5	Is this a change to sterilisation?	<i>Not applicable, Leica Biosystems Melbourne does not manufacture sterile products.</i>	
6.A6	Is this a change to software?	☐ Yes <i>Go to 6.E</i>	☒ No
	<i>Detailed justification:</i> No change to software.		
6.A7	Is this a change in materials for an IVDD?	☐ Yes <i>Go to 6.F</i>	☒ No
	<i>Detailed justification:</i> There are no changes to the material of an IVDD; the refurbished instruments will be built using the same material specifications as those of newly manufactured instruments.		
6.A8	Is this a change to labelling?	☒ Yes <i>Go to 6.G</i>	☐ No
	<i>Detailed justification:</i> Two new crate label and two rating plate labels are being created as part of this change. The content/symbols of the label remain the same compared to the one of a new instrument.		
6.A9	Is the safety and effectiveness of the device affected?	☐ Yes <i>Amendment required</i>	☒ No
	<i>Detailed justification:</i> No impact to the functionality of the instrument. No impact to safety and effectiveness.		

REGULATORY CHANGE ASSESSMENT FORM

6.A10	Has this change been notified to the Medical Devices Bureau previously under MDR Section 34(e)? <i>NOTE: Submit changes with license renewal only if answers to 6.A1 – 6.A9 are all NO]</i>	☐ Yes <i>Amendment not required</i>	☒ No <i>Submit changes with license renewal</i>
	<i>Detailed justification:</i> This change will not be submitted with license renewal as the answer to 6.A8 is "Yes".		

REGULATORY CHANGE ASSESSMENT FORM

6.B	Is this a change to manufacturing processes, facility or equipment?	☐ Yes	☒ No Go to 6.C
6.B1	Does the change impact the device safety or effectiveness?	☐ Yes <i>Amendment required</i>	☐ No
<i>Detailed justification:</i>			
6.B2	Is the manufacturer's name and address on the device labelling staying the same, but a new manufacturing facility is added?	☐ Yes <i>Amendment required</i>	☐ No
<i>Note – if the answer is YES, the new facility will need to be covered by the manufacturer's quality system certificate.</i>			
<i>Detailed justification:</i>			
6.B3	Has there been a change in a supplier's manufacturing process, facility or equipment?	☐ Yes	☐ No
<i>Note – an amendment is not required provided device specifications have not changed and incoming inspections to evaluate the material/equipment provided by the supplier have not changed.</i>			
<i>Detailed justification:</i>			
6.B4	Is it a change to a combination product?	☐ Yes	☐ No
<i>Note – if the answer is YES, consult the policy Drug/Medical Device Combination Products or call the Medical Devices Bureau for clarification.</i>			
<i>Detailed justification:</i>			
6.B5	Is it a change to sterilisation method or procedures?	☐ Yes Go to 6.E	☐ No
<i>Detailed justification:</i>			
6.C	Is this a change to manufacturing control procedures?	☐ Yes	☒ No Go to 6.D
6.C1	Is the change due to removal of test acceptance criteria, in process inspections, or final inspections without replacement of these activities?	☐ Yes <i>Amendment required</i>	☐ No
<i>Detailed justification:</i>			
6.C2	Does the change alter the design specifications of the device?	☐ Yes Go to 6.D	☐ No
<i>Detailed justification:</i>			

REGULATORY CHANGE ASSESSMENT FORM

6.D	Is this a change to design?	☐ Yes	☒ No Go to 6.E
6.D1	Is this a change to the control mechanism (mechanism put in place to maintain on-going control or regulate the output of a device)?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.D2	Is this a change to the operating principles, including a change in the source of energy used by the device?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.D3	Do the changes to the design specification, physical description, patient or user interface, software or firmware affect the indications for use?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.D4	Are clinical data necessary to support the safety and effectiveness of the altered device?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.D5	Do the results of a risk analysis, undertaken during the design validation process, raise new issues of safety and effectiveness?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		

REGULATORY CHANGE ASSESSMENT FORM

6.E	Is this a software change?	☐ Yes	☒ No Go to 6.F
6.E1	Does the change impact the control of the device in a manner that might alter the diagnosis or therapy delivered?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E2	Does the change in software modify an algorithm that impacts the diagnosis or therapy delivered?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E3	Does the change in software impact the way data is read and/or interpreted, such that the treatment or diagnosis of the patient may be altered when compared to previous version of the software?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E4	Does the change in software modify a decision point from requiring user input to fully automated (e.g. closed loop decision)?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E5	Does the change in software add a new feature that may change the diagnosis or therapy delivered?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E6	Does the change in software introduce or remove an alarm function, and a response to the new alarm may change the treatment of the patient in comparison to the previous version of the software?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E7	Does the change in software incorporate a change to the operating system on which the software is running?	☐ Yes <i>Amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E8	Does the change correct an error for which there is a safety risk to the patient if the error is not fixed?	☐ Yes <i>Contact HC for clarification</i>	☐ No
	<i>Detailed justification:</i>		

REGULATORY CHANGE ASSESSMENT FORM

6.E9	Does the change in software correct an inadvertent logic error to bring the system back into specification?	☐ Yes <i>No amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E10	Does the change in software only introduce non-therapeutic and non-diagnostic features (e.g. printing, faxing, improved image clarity or reporting format)?	☐ Yes <i>No amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E11	Does the change in software only modify the user interface in appearance with negligible risk of impacting diagnosis or therapy delivered?	☐ Yes <i>No amendment required</i>	☐ No
	<i>Detailed justification:</i>		
6.E12	Does the change in software disable a feature that does not interact with other features?	☐ Yes <i>No amendment required</i>	☐ No <i>Contact HC for clarification (only if answers to 6.E1 – 6.E11 are “No”)</i>
	<i>Detailed justification:</i>		

REGULATORY CHANGE ASSESSMENT FORM

6.F	Is this a change to materials for an IVDD?	☐ Yes	☒ No Go to 6.G
6.F1	Does the change necessitate the testing of additional clinical samples to determine the performance characteristics of the IVDD?	☐ Yes	☐ No Go to 6.G
	<i>Detailed justification</i>		
6.F2	Does the additional clinical testing completed confirm that the altered IVDD still conforms to the licensed performance specifications and no labelling changes are necessary?	☐ Yes	☐ No Amendment required
	<i>Detailed justification:</i>		
6.G	Is this a change to labelling?	☒ Yes	☐ No
6.G1	Is there a change in the indications for use that are not within an approved set of indications?	☐ Yes Amendment required	☒ No
	<i>Detailed justification:</i> There is no change in the indications for use that are not within an approved set of indications.		
6.G2	Does the change to the existing warnings and precautions add or remove a contraindication?	☐ Yes Amendment required	☒ No
	<i>Detailed justification:</i> There is no change to the existing warnings and precautions add or remove a contraindication		
6.G3	Does the change include the deletion of a warning or precaution?	☐ Yes Amendment required	☒ No
	<i>Detailed justification:</i> There is no change to precaution or warnings		
6.G4	Is there a change in the shelf life of an in vitro diagnostic device?	☐ Yes Amendment required	☒ No
	<i>Detailed justification:</i> There is no change in the shelf life of an in vitro diagnostic device		

REGULATORY CHANGE ASSESSMENT FORM

SECTION 8: Assessment for Japan registration impact

8.A	Is this a change to a Japan Class I device?	☒ Yes	☐ No Go to 9
8.A1	The changes listed below are examples of changes that may require partial change to the approved information: • Intended purpose or other changes that affect the safety or efficacy of the device (e.g. changes to the operating principle, materials or user interface). • Will the change increase the foreseeable risks? • Will the change produce new risks for patients? • Do the changes in structure and principle lead to the IVD being different from the parent IVD?	☐ Yes Partial change	☒ No
	<i>Detailed justification:</i> This change will add two new codes for refurbished instruments without changing the device itself. No change in product's classification or intended use and no change to safety and effectiveness.		
8.A2	Are these minor changes in compliance with design change requirements of the quality management system?	☒ Yes No regulatory procedure is needed	☐ No
	<i>Detailed justification:</i> This change will add two new codes for refurbished instruments without changing the device itself. No change in product's classification or intended use and no change to safety and effectiveness.		
8.A3	Is this a change to the JMDN code?	☐ Yes Notify local regulatory affairs unit for assessment	☒ No
	<i>Detailed justification:</i> No change to the JMDN code.		

REGULATORY CHANGE ASSESSMENT FORM

SECTION 9: Assessment for Australia registration impact

9.A	Is this a change to an Australia Class I device?	☒ Yes	☐ No <i>Go to 10</i>
9.A1	The changes listed below are examples of changes that need to be notified to the TGA: • Change to add, modify or remove the legal manufacture's name or address. • Change in the intended purpose of the device. • Change in the risk classification. • Significant change to the design or composition of the device that may affect its safety or performance. • Change that adds or deletes an IVD or accessory in the AU Declaration of Conformity.	☒ Yes <i>Approval required</i>	☐ No
<i>Detailed justification:</i> This change will add two new codes for refurbished instruments without changing the device itself. No change in product's classification or intended use and no change to safety and effectiveness.			

REGULATORY CHANGE ASSESSMENT FORM

SECTION 10: Assessment for rest of world (ROW) registration impact

10.A	Is it a design or technology change?	☐ Yes	☒ No <i>Go to 10.B</i>
10.A1	Does the change introduce new hazards which have not been previously addressed? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A2	Does the change adversely affect the risk associated with existing hazards? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A3	Does the change significantly alter performance characteristics or product claims as stated in the product labelling? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A4	Does the change alter the testing methods of the product (change to operating principle, control mechanism or source of energy)? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A5	Does the change affect product conformity to essential principles, General Safety and Performance Requirements and/or Canada Safety and Effectiveness and Labelling requirements? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A6	Does the change cause the device to be a different classification? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A7	Does the change affect stability or expiration claims? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A8	Is there a change to software operating platform or core algorithm? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A9	Is it a change to the published software version (IFU or UI screen)? <i>Detailed justification:</i>	☐ Yes	☐ No
10.A10	Does the change require further clinical or usability data to support safety and performance? <i>Detailed justification:</i>	☐ Yes	☐ No

REGULATORY CHANGE ASSESSMENT FORM

10.B	Is it a materials/manufacturing change?	☐ Yes	☒ No Go to 10.C
10.B1	Is it a change to a critical supplier or material that impacts product safety or performance? <i>Detailed justification:</i>	☐ Yes	☐ No
10.B2	Does the change significantly alter the manufacturing process, manufacturing method, QC procedures, or product release specifications? <i>Detailed justification:</i>	☐ Yes	☐ No
10.B3	Is it a change to a supplier of an accessory listed on the EU Declaration of Conformity? <i>Detailed justification:</i>	☐ Yes	☐ No
10.B4	Does a risk assessment of the changed device identify any new risks or significantly modified existing risks? <i>Detailed justification:</i>	☐ Yes	☐ No
10.B5	Is there a manufacturing site change? <i>Detailed justification:</i>	☐ Yes	☐ No

10.C	Is it a labelling change?	☒ Yes	☐ No Go to 10.D
10.C1	Does the change alter the intended use and/or indications for use? <i>Detailed justification:</i> No changes to the intended use, not impact to the function of the instrument or how the device is used.	☐ Yes	☒ No
10.C2	Does the change affect the product name or alter the product part number? <i>Detailed justification:</i> This change will add two new codes for refurbished instruments without changing the device itself or product name. No change in product's classification or intended use and no change to safety and effectiveness.	☒ Yes	☐ No
10.C3	Does the change add or alter a sample type? <i>Detailed justification:</i> No change to add or alter a sample type.	☐ Yes	☒ No
10.C4	Does the change add or remove a warning or contraindication? <i>Detailed justification:</i> No change to add or remove a warning or contraindication.	☐ Yes	☒ No
10.C5	New user or patient population? <i>Detailed justification:</i> No change to the user or patient population.	☐ Yes	☒ No
10.C6	Does this change impact the Country of Origin on the label? <i>Detailed justification:</i> This is not a change to the Country of Origin information on the labels.	☐ Yes	☒ No

REGULATORY CHANGE ASSESSMENT FORM

10.D	Is this a change to the quality system?	☐ Yes	☒ No Go to 10.E
10.D1	Is it a change to the name and/or address of the manufacturer?	☐ Yes	☐ No
<i>Detailed justification:</i>			
10.D2	Is it a change to the product range covered by the quality system?	☐ Yes	☐ No
<i>Detailed justification:</i>			
10.D3	Does the change affect the continued compliance of the quality system with the relevant standards?	☐ Yes	☐ No
<i>Detailed justification:</i>			

10.E	EU Technical Documentation Update?	☒ Yes	☐ No Go to 10.F
10.E1	Are there any changes to product technical documentation as a result of this change?	☒ Yes	☐ No
<i>If Yes, update relevant technical documentation and provide copies to EU Authorised Representative.</i>			
<i>Detailed justification:</i> The relevant updated technical documentation will be submitted to EU Authorised Representative.			

10.F	Is this a change to the Unique Device Identifier (UDI)?		
10.F1	Does this change impact the UDI-DI for the product and accessories as identified in the Declaration of Conformity?	☒ Yes	☐ No
<i>Consider if the information registered in applicable UDI database requires updating, for example:</i> • Product name change • New model or catalogue number • Change to critical warnings or contraindications • Change to quantity of devices in the packaging • Change in product colour/language which might lead to misidentification of product • New device or accessory		<i>If yes, a new UDI-DI is required.</i>	
<i>Detailed justification:</i> UDI-DI information was updated, since this change is to generate new part numbers for the refurbished instruments.			

10.G	Summary		
10.G1	Do any product registrations require updating due to this change?	☒ Yes	☐ No
<i>NOTE: If answers to any of the questions in Section 7 are YES then the change may be considered significant and may require notification to a regulatory authority or local regulatory affairs unit.</i>			

REGULATORY CHANGE ASSESSMENT FORM

Detailed justification: Registration activities for the two new product codes might be required in the countries where BOND is registered (a current list is detailed in Appendix A). Products will be supplied after receiving approval from the relevant regulatory authority only (where applicable). Shipment of the new refurbished products will be restricted via Licence Control in SAP in accordance with MEL-WI-000079 Licence Control in SAP until confirmation of registration, where applicable, has been received. Refer to Appendix A for details.

SECTION 11: Regulatory Change Assessment Summary

11.1 Applicable Regulatory History Summary and Conclusion:

USA: This change does not impact US registrations:

U.S 510 (k) #: K183100,

U.S 510 (k) #: K193393,

U.S 510 (k) #: K213348,

U.S PMA #: P090015

Canada: This change does not impact Canadian registrations:

Device license # 81061, Class II (BOND-III System)

Device license # 89628, Class III (BOND-III System)

Device license # 66466, (BOND-MAX System)

Device license # 89629, Class III (BOND-MAX System)

Rest of World:

As part of this change registration activities for the two new product codes might be required in the countries where BOND is registered (a current list is detailed in Appendix A).

Australia – BOND is a Class 1 class I IVD included on the ARTG entry 176143.

Japan – This change does not impact the Japanese registration:

Device license # 13B2X10268BM0001(BOND-MAX)

Device license # 13B2X10268B30001 (BOND-III)

11.2 Registration Impact Assessment:

510 (k) Assessment (Applicable/ ~~Not applicable~~)

☐ Submit 510(k)

☒ Letter to File

PMA Assessment (Applicable/ ~~Not applicable~~)

☐ 180-day PMA Supplement

☐ 30 Day Notice

☐ Real-time Supplement

☒ Annual Report

☐ Special PMA Supplement

510(k) Exempt Assessment (Applicable/ ~~Not applicable~~)

☐ Administrative update to FURLS

☒ Document as part of change control

☐ Submit 510(k)

Canada Class II Assessment (Applicable/ ~~Not applicable~~)

☒ License amendment

☒ Document as part of change control

Canada Class III/IV Assessment (Applicable/ ~~Not applicable~~)

☒ License amendment

☒ Document as part of change control

☐ Submit changes with license renewal

REGULATORY CHANGE ASSESSMENT FORM

Brazil Class I Assessment (Applicable / Not applicable)	Japan Class I Assessment (Applicable / Not applicable)
☐ Approval required by regulatory authority ☐ Notify regulatory authority ☐ Document as part of change control	☐ Approval required by regulatory authority ☒ Notify local regulatory affairs unit for assessment ☒ Document as part of change control
Australia Class 1 Assessment (Applicable/ Not applicable)	ROW Assessment (Applicable/ Not applicable)
☒ Approval required by regulatory authority ☒ Document as part of change control	☒ Notify regulatory authority ☒ Notify local regulatory affairs unit for assessment ☒ Notify EU authorised representative (Update to Technical Documentation) ☒ Document as part of change control
UDI Assessment (Applicable/ Not applicable)	
Administrative update to UDI database: ☒ Yes Database name(s): GUDID ☐ No	
11.3 Supporting Documents:	
DHFR_18528_A01 Impact Assessment BOND Refurbished CR12773	

Appendix A

Countries in which BOND is currently available for sale and will be eligible to receive the Refurbished BOND-III and/or BOND-MAX instruments once registration is completed with the relevant regulatory authorities (where applicable, in accordance with MEL-SOP-000064):

Australia	UAE
Bosnia	UK
Botswana	Ukraine
Canada	Uzbekistan
Egypt	
Ghana	
Greece	
Israel	
Italy	
Japan	
Jordan	
Kazakhstan	
Kenya	
Kuwait	
Macedonia	
Moldova	
Montenegro	
Nigeria	
Oman	
Portugal	
Russia	
Saudi Arabia	
Serbia	
South Africa	
Tanzania	
Turkey	