PanGuide 3D

Experimental segmentation | research and education only


Segmentation summary (mask geometry only)


Predicted pancreas	79.46 cm3
Predicted tumor	8.23 cm3
Intersection	4.56 cm3
Connected tumor regions	2

Tumor localization and review support
Recorded Demo Results | cached from a completed real model run
Examination overview | CT | original dimensions (X,Y,Z): (512, 512, 41) | 41 slices | spacing (0.816, 0.816, 5.0) mm.
Pancreas localization | predicted mask volume 79.46 cm3.
Candidate lesion segmentation | 8.23 cm3 | 2 connected predicted regions.
Geometric intersection | 4.56 cm3 | 55.4% of tumor mask inside pancreas mask | 5.7% of pancreas mask intersected. Mathematical overlap is not evidence of malignancy.
Tumor segmentation confidence
Mean model score across the predicted tumor mask: 91.2%
Uncalibrated model score; not the probability that the region is cancerous. High scores can still be wrong.
Region (largest first)	Mean model score	Voxels
1	91.8%	2,296
2	82.9%	174
Scores average the network's tumor-channel sigmoid outputs, linearly resampled onto the original CT grid, within each displayed six-connected region. They describe the model's label preference, not diagnostic accuracy or a calibrated clinical risk.
Scan-specific explanation
The model marked 2 candidate tumor region(s), totaling 8.23 cm3 of predicted mask volume. Their mathematical intersection with the predicted pancreas is 4.56 cm3 (55.4% of the candidate tumor mask). These views support inspection of location and spatial relationships during research review. The highlighted tissue is not confirmed cancer; this output cannot determine tumor type, stage, invasion, or treatment.
Model: PanGuide3D, 52,328,834 parameters | checkpoint 917acfa35774 | web software 0.3.0.
Runtime: 109.837 seconds. Peak CUDA allocated memory: not measured.

PanGuide 3D | predicted overlays | educational use only




S
R
L
I


S
A
P
I