
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001164
10:11 12/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:38
STERILIZING 5:00
136.0 C 220kPa 5:00
136.4 C 224kPa 5:30
136.4 C 225kPa 6:00
136.4 C 225kPa 6:30
136.3 C 224kPa 7:00
136.3 C 223kPa 7:30
136.4 C 225kPa 8:00
136.3 C 224kPa 8:30
136.3 C 224kPa 8:30
VENTING 8:30
AIR DRYING 9:24
CYCLE COMPLETE 22:09

Digital Signature #
1463A7188E656BDF
