

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3679

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2024 03:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.8 F 0.0  
2:00 138.1 F 1.9  
4:00 149.6 F 4.9  
6:00 176.3 F 8.7  
8:00 195.6 F 13.2  
10:00 215.8 F 6.8  
12:00 227.8 F 11.5  
14:00 229.5 F 8.1  
16:00 241.2 F 12.5  
18:00 242.4 F 12.5  
20:00 254.5 F 18.7  
22:00 265.3 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.5 F 30.2  
2:00 271.7 F 29.6  
3:00 271.7 F 30.0  
4:00 272.5 F 30.1

Min 271.1 F 29.4  
Max 272.7 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 211.9 F 0.0  
10:00 212.1 F 0.0  
15:00 215.7 F 0.0  
20:00 213.7 F 0.0  
25:00 214.8 F 0.0  
30:00 211.5 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 23:51  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:02:28  
CYCLE COMPLETE

06/07/2024 04:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_