
Midmark M11 - v1.0.5
Total Cycles: 2968

Name: RITTER2
Sterilizer ID: V2313228
10/04/2023 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.6 F 0.0
2:00 154.8 F 2.1
4:00 168.8 F 5.5
6:00 194.0 F 9.7
8:00 214.0 F 8.1
10:00 226.2 F 8.8
12:00 233.3 F 11.6
14:00 237.9 F 11.4
16:00 238.0 F 10.0
18:00 251.1 F 16.2
20:00 261.6 F 22.6
22:00 271.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 271.6 F 29.3
2:00 271.5 F 29.2
3:00 271.5 F 29.3
4:00 272.0 F 29.5

Min 271.3 F 29.2
Max 272.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 222.2 F 0.0
10:00 218.7 F 0.0
15:00 200.3 F 0.0
20:00 198.3 F 0.0
25:00 210.9 F 0.0
30:00 213.6 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

10/04/2023 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____