
Midmark M11 - v1.0.5
Total Cycles: 1881

Name: RITTER2
Sterilizer ID: V2313228
12/22/2022 12:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.4 F 0.0
2:00 180.8 F 3.3
4:00 184.6 F 8.2
6:00 204.8 F 10.1
8:00 225.2 F 8.8
10:00 236.9 F 10.2
12:00 241.0 F 12.5
14:00 241.3 F 11.5
16:00 254.7 F 18.5
18:00 266.6 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.7
2:00 271.7 F 28.9
3:00 272.1 F 29.7
4:00 272.4 F 29.5

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 213.2 F 0.0
10:00 223.9 F 0.0
15:00 228.1 F 0.0
20:00 228.5 F 0.0
25:00 228.4 F 0.0
30:00 227.5 F 0.0

Drying Complete

FILLING: 2:05
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

12/22/2022 01:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____