
Midmark M11 - v1.0.0
Total Cycles: 4542

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2024 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.5 F 0.0
2:00 138.7 F 1.8
4:00 152.1 F 4.9
6:00 174.4 F 8.3
8:00 200.6 F 12.4
10:00 218.7 F 5.4
12:00 230.0 F 9.4
14:00 239.4 F 12.6
16:00 237.3 F 9.7
18:00 239.1 F 9.9
20:00 245.3 F 12.8
22:00 256.2 F 18.6
24:00 266.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.7 F 29.1
2:00 271.7 F 29.0
3:00 271.4 F 28.5
4:00 272.7 F 28.9

Min 270.6 F 28.1
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 216.5 F 0.0
10:00 214.8 F 0.0
15:00 199.3 F 0.0
20:00 193.6 F 0.0
25:00 193.7 F 0.0
30:00 190.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:04:11
CYCLE COMPLETE

05/10/2024 12:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____