
Midmark M11 - v1.0.5
Total Cycles: 4500

Name: RITTER2
Sterilizer ID: V2313228
04/09/2025 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.7 F 0.0
2:00 101.3 F 0.6
4:00 104.3 F 2.1
6:00 146.1 F 5.0
8:00 158.6 F 8.5
10:00 179.7 F 12.3
12:00 218.3 F 4.7
14:00 226.0 F 8.4
16:00 234.7 F 12.4
18:00 232.5 F 8.2
20:00 243.3 F 13.1
22:00 241.0 F 11.2
24:00 252.5 F 17.0
26:00 262.8 F 23.0
28:00 272.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.3
1:00 273.0 F 29.5
2:00 272.6 F 29.7
3:00 272.1 F 29.9
4:00 271.2 F 29.6

Min 271.2 F 29.2
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 227.9 F 0.0
10:00 171.3 F 0.0
15:00 164.2 F 0.0
20:00 161.7 F 0.0
25:00 158.9 F 0.0
30:00 158.3 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:58
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:07:41
CYCLE COMPLETE

04/09/2025 04:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____