
Midmark M11 - v1.0.5
Total Cycles: 3008

Name: RITTER2
Sterilizer ID: V2313228
10/13/2023 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.4 F 0.0
2:00 193.2 F 4.1
4:00 210.7 F 10.0
6:00 224.0 F 6.8
8:00 236.3 F 11.7
10:00 236.7 F 10.0
12:00 236.3 F 9.0
14:00 251.4 F 16.4
16:00 263.7 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 272.1 F 29.6
2:00 272.3 F 29.2
3:00 271.4 F 29.0
4:00 272.7 F 29.7

Min 271.3 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 220.5 F 0.0
10:00 218.9 F 0.0
15:00 221.0 F 0.0
20:00 220.0 F 0.0
25:00 216.2 F 0.0
30:00 213.9 F 0.0

Drying Complete

FILLING: 2:25
HEATING: 17:53
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

10/13/2023 12:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____