
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001919
10:23 07/08/2024
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 4:15
STERILIZING 5:10
136.0 C 222kPa 5:10
136.1 C 223kPa 5:40
136.4 C 226kPa 6:10
136.3 C 225kPa 6:40
136.3 C 225kPa 7:10
136.3 C 225kPa 7:40
136.3 C 225kPa 8:10
136.4 C 226kPa 8:40
136.4 C 225kPa 9:10
136.4 C 223kPa 9:40
136.3 C 224kPa 10:10
136.4 C 225kPa 10:40
136.2 C 224kPa 11:10
136.4 C 225kPa 11:40
136.4 C 224kPa 12:10
136.4 C 223kPa 12:40
136.2 C 224kPa 13:10
136.4 C 226kPa 13:40
136.4 C 226kPa 14:10
136.4 C 226kPa 14:40
136.2 C 224kPa 15:10
136.2 C 224kPa 15:10
VENTING 15:11
AIR DRYING 16:03
CYCLE COMPLETE 46:03

Digital Signature #
A5D60F73F36E3DF3
