
Midmark M11 - v1.0.5
Total Cycles: 2445

Name: RITTER2
Sterilizer ID: V2313228
05/31/2023 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.0 F 0.0
2:00 151.1 F 0.0
4:00 188.5 F 3.1
6:00 206.2 F 7.8
8:00 211.9 F 12.9
10:00 224.5 F 7.3
12:00 237.4 F 12.3
14:00 235.0 F 9.5
16:00 238.9 F 9.5
18:00 247.5 F 14.5
20:00 259.6 F 21.2
22:00 270.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.7 F 29.6
2:00 271.7 F 29.2
3:00 272.7 F 29.7
4:00 271.8 F 29.0

Min 271.6 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 209.5 F 0.0
10:00 229.5 F 0.0
15:00 236.9 F 0.0
20:00 227.0 F 0.0
25:00 226.9 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

05/31/2023 02:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____