
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007603
9:59 27/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:52
STERILIZING 3:59
136.0 C 220kPa 3:59
136.1 C 221kPa 4:29
136.0 C 221kPa 4:59
135.9 C 221kPa 5:29
136.2 C 223kPa 5:59
136.2 C 221kPa 6:29
136.4 C 224kPa 6:59
136.2 C 219kPa 7:29
136.2 C 219kPa 7:29
VENTING 7:29
AIR DRYING 8:18
CYCLE COMPLETE 8:28

Digital Signature #
5A2F43260B0D6908
