

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4197

Name: RITTER2  
Sterilizer ID: V2313228  
12/09/2024 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 71.4 F 0.0  
4:00 89.3 F 0.0  
6:00 117.6 F 0.2  
8:00 158.2 F 0.0  
10:00 185.5 F 2.4  
12:00 201.5 F 6.4  
14:00 215.8 F 10.1  
16:00 223.2 F 8.6  
18:00 232.6 F 9.5  
20:00 237.9 F 10.2  
22:00 242.7 F 12.2  
24:00 241.1 F 11.0  
26:00 255.2 F 18.1  
28:00 267.2 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.4 F 28.8  
2:00 272.7 F 29.3  
3:00 271.6 F 28.4  
4:00 272.8 F 29.1

Min 271.5 F 28.3  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.0 F 0.0  
Dry Cycle Aborted

FILLING: 0:57  
HEATING: 29:21  
STERILIZING: 4:00  
VENTING: 2:38  
DRYING: 0:38  
TOTAL CYCLE: 00:37:35

12/09/2024 11:38 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_