
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000894
15:56 31/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:45
STERILIZING 4:53
136.0 C 228kPa 4:53
136.1 C 230kPa 5:23
136.4 C 233kPa 5:53
136.4 C 231kPa 6:23
136.3 C 232kPa 6:53
136.1 C 230kPa 7:23
136.5 C 233kPa 7:53
136.4 C 233kPa 8:23
136.4 C 233kPa 8:23
VENTING 8:23
AIR DRYING 9:18
CYCLE COMPLETE 19:53

Digital Signature #
BC3C646FE1A2D5D4
