
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002058
12:50 19/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:52
STERILIZING 4:00
136.4 C 223kPa 4:00
136.3 C 224kPa 4:30
136.4 C 222kPa 5:00
136.4 C 224kPa 5:30
136.0 C 221kPa 6:00
136.1 C 222kPa 6:30
136.4 C 224kPa 7:00
136.1 C 222kPa 7:30
136.1 C 222kPa 7:30
VENTING 7:30
AIR DRYING 8:25
CYCLE COMPLETE 38:25

Digital Signature #
C293B4820F127FEF
