
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001399
8:35 09/08/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 6:04
STERILIZING 7:32
136.5 C 222kPa 7:32
136.3 C 222kPa 8:02
136.2 C 221kPa 8:32
136.4 C 223kPa 9:02
136.2 C 221kPa 9:32
136.2 C 220kPa 10:02
136.4 C 221kPa 10:32
136.3 C 221kPa 11:02
VENTING 11:02
AIR DRYING 12:03
CYCLE COMPLETE 14:11

Digital Signature #
BD777C3A730795A5
