
Midmark M11 - v1.0.5
Total Cycles: 3526

Name: RITTER2
Sterilizer ID: V2313228
04/15/2024 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.9 F 0.0
2:00 126.5 F 0.2
4:00 130.9 F 2.1
6:00 169.9 F 5.7
8:00 191.1 F 9.9
10:00 230.3 F 8.2
12:00 228.0 F 8.1
14:00 238.1 F 12.7
16:00 237.0 F 9.5
18:00 238.8 F 9.9
20:00 246.6 F 13.8
22:00 258.0 F 20.1
24:00 268.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 271.6 F 29.3
2:00 272.6 F 29.5
3:00 272.5 F 29.7
4:00 272.0 F 29.9

Min 271.0 F 29.3
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.8 F 0.0
5:00 210.1 F 0.0
10:00 189.3 F 0.0
15:00 188.5 F 0.0
20:00 188.4 F 0.0
25:00 188.2 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:55
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:04:26
CYCLE COMPLETE

04/15/2024 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____