
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000251
10:22 21/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:58
PRESSURIZING 3:58
STERILIZING 5:00
136.5 C 224kPa 5:00
136.4 C 225kPa 5:31
136.3 C 222kPa 6:01
136.4 C 221kPa 6:31
136.4 C 221kPa 7:01
136.3 C 222kPa 7:31
136.4 C 224kPa 8:01
136.3 C 223kPa 8:31
136.3 C 223kPa 8:31
VENTING 8:31
AIR DRYING 9:24
CYCLE COMPLETE 9:45

Digital Signature #
F9DE21B89F0A3AF5
