

-----  
Midmark M11 - v1.0.5  
Total Cycles: 908

Name: RITTER2  
Sterilizer ID: V2313228  
03/25/2022 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.9 F 0.0  
2:00 151.6 F 0.2  
4:00 183.9 F 2.6  
6:00 205.5 F 7.3  
8:00 231.7 F 12.3  
10:00 228.9 F 7.7  
12:00 241.0 F 13.1  
14:00 240.4 F 11.4  
16:00 239.9 F 10.8  
18:00 253.5 F 17.7  
20:00 265.4 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.8 F 29.7  
2:00 271.5 F 28.9  
3:00 272.6 F 29.8  
4:00 272.6 F 29.3  
  
Min 271.3 F 28.8  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.0 F 0.0  
5:00 228.7 F 0.0  
10:00 220.1 F 0.0  
15:00 220.4 F 0.0  
20:00 218.1 F 0.0  
25:00 212.6 F 0.0  
30:00 216.8 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 21:47  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:00:21  
CYCLE COMPLETE

03/25/2022 01:32 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02