
Midmark M11 - v1.0.0
Total Cycles: 1538

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2022 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.0 F 0.0
2:00 173.4 F 3.5
4:00 185.8 F 7.9
6:00 213.1 F 12.5
8:00 221.1 F 7.9
10:00 235.0 F 13.2
12:00 237.3 F 10.5
14:00 235.6 F 8.8
16:00 250.5 F 15.9
18:00 262.7 F 22.8
20:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.4 F 29.1
2:00 271.5 F 28.9
3:00 272.5 F 29.1
4:00 271.3 F 28.6

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 234.0 F 0.0
10:00 239.6 F 0.0
15:00 232.3 F 0.0
20:00 225.9 F 0.0
25:00 227.6 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

01/19/2022 01:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____